



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110223\
Data File : VD077485.D
Acq On : 02 Nov 2023 09:09
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Title : SW846 8260
Last Update : Tue Oct 31 04:19:17 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.0	36544	PASS
75	95	30	60	56.8	86621	PASS
95	95	100	100	100.0	152427	PASS
96	95	5	9	6.7	10250	PASS
173	174	0.00	2	1.6	1797	PASS
174	95	50	100	75.0	114384	PASS
175	174	5	9	7.9	9007	PASS
176	174	95	101	99.7	114093	PASS
177	176	5	9	7.0	7960	PASS