

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
 Data File : VD074733.D
 Acq On : 04 Nov 2022 12:59
 Operator : VA/SY
 Sample : N5430-03
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-COMP-1102

Quant Time: Nov 07 00:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	150291	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	234390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	207776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	80416	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	75961	63.815	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.620%
35) Dibromofluoromethane	7.805	113	93619	67.521	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	135.040%
50) Toluene-d8	10.269	98	282370	55.831	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.660%
62) 4-Bromofluorobenzene	12.575	95	77027	52.748	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
Data File : VD074733.D
Acq On : 04 Nov 2022 12:59
Operator : VA/SY
Sample : N5430-03
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-COMP-1102

Quant Time: Nov 07 00:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

