

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
 Data File : VD074747.D
 Acq On : 04 Nov 2022 19:28
 Operator : VA/SY
 Sample : N5433-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-01-110222

Quant Time: Nov 07 00:29:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	7694	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	12615	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	14442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7191	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	8631	141.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	283.280%#
35) Dibromofluoromethane	7.804	113	7539	101.028	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	202.060%#
50) Toluene-d8	10.269	98	16236	59.647	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.300%
62) 4-Bromofluorobenzene	12.575	95	6653	84.651	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	169.300%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
Data File : VD074747.D
Acq On : 04 Nov 2022 19:28
Operator : VA/SY
Sample : N5433-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-01-110222

Quant Time: Nov 07 00:29:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

