

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
 Data File : VD074742.D
 Acq On : 04 Nov 2022 17:09
 Operator : VA/SY
 Sample : N5418-04
 Misc : 3.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-K

Quant Time: Nov 07 00:28:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	122184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	229571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	191722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	87340	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	71474	73.858	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.720%
35) Dibromofluoromethane	7.804	113	84195	61.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	124.000%
50) Toluene-d8	10.269	98	248899	50.246	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.575	95	76262	53.320	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110422\
Data File : VD074742.D
Acq On : 04 Nov 2022 17:09
Operator : VA/SY
Sample : N5418-04
Misc : 3.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-K

Quant Time: Nov 07 00:28:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09




Instrument :
MSVOA_D
ClientSampleId :
TP-K


Tgt Ion:168 Resp: 122184
Ion Ratio Lower Upper
168 100
99 52.0 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 73.858 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09

Tgt Ion: 65 Resp: 71474
Ion Ratio Lower Upper
65 100
67 55.2 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074742.D

Acq: 04 Nov 2022 17:09

Instrument :

MSVOA_D

ClientSampleId :

TP-K





Tgt Ion:114 Resp: 229571

Ion Ratio Lower Upper

114 100

63 17.8 0.0 33.2

88 13.6 0.0 28.2



#35

Dibromofluoromethane

Concen: 61.999 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.001 min

Lab File: VD074742.D

Acq: 04 Nov 2022 17:09




Tgt Ion:113 Resp: 84195


Ion Ratio Lower Upper

113 100

111 102.7 82.3 123.5

192 20.0 16.4 24.6









#50
Toluene-d8
Concen: 50.246 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09




Instrument :
MSVOA_D
ClientSampleId :
TP-K


Tgt Ion: 98 Resp: 248899
Ion Ratio Lower Upper
98 100
100 63.7 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 53.320 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09

Tgt Ion: 95 Resp: 76262
Ion Ratio Lower Upper
95 100
174 95.6 0.0 197.0
176 91.8 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09

Instrument :
MSVOA_D
ClientSampleId :
TP-K

Tgt Ion:117 Resp: 191722
Ion Ratio Lower Upper
117 100
82 51.3 41.6 62.4
119 33.0 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD074742.D
Acq: 04 Nov 2022 17:09

Tgt Ion:152 Resp: 87340
Ion Ratio Lower Upper
152 100
115 54.0 26.8 80.4
150 157.7 0.0 348.2

