

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110623\
 Data File : VD077507.D
 Acq On : 06 Nov 2023 12:01
 Operator : JC/SY
 Sample : VD1106SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Nov 07 02:08:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	223882	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	397362	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	349930	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	160722	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	145496	45.959	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.920%
35) Dibromofluoromethane	7.804	113	128407	46.754	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.500%
50) Toluene-d8	10.269	98	533339	49.995	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.000%
62) 4-Bromofluorobenzene	12.575	95	166505	46.342	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.680%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.934	85	34868	19.315	ug/l	98
3) Chloromethane	2.152	50	53011	18.496	ug/l	99
4) Vinyl Chloride	2.287	62	62065	22.207	ug/l	99
5) Bromomethane	2.687	94	35029	23.919	ug/l	97
6) Chloroethane	2.840	64	44126	23.451	ug/l	98
7) Trichlorofluoromethane	3.175	101	74517	19.039	ug/l	97
8) Diethyl Ether	3.593	74	25687	18.258	ug/l #	55
9) 1,1,2-Trichlorotrifluo...	3.969	101	47031	19.905	ug/l	94
10) Methyl Iodide	4.163	142	44644	17.167	ug/l	92
11) Tert butyl alcohol	5.052	59	23213	92.566	ug/l	98
12) 1,1-Dichloroethene	3.940	96	40477	18.095	ug/l #	72
13) Acrolein	3.799	56	30634	92.635	ug/l	97
14) Allyl chloride	4.558	41	96337	17.853	ug/l #	84
15) Acrylonitrile	5.258	53	75051	98.683	ug/l	99
16) Acetone	4.022	43	86475	103.228	ug/l #	81
17) Carbon Disulfide	4.275	76	79760	14.448	ug/l	95
18) Methyl Acetate	4.563	43	54854	17.952	ug/l #	79
19) Methyl tert-butyl Ether	5.316	73	132150	18.821	ug/l	96
20) Methylene Chloride	4.805	84	73469	21.974	ug/l #	62
21) trans-1,2-Dichloroethene	5.316	96	46192	17.731	ug/l #	78
22) Diisopropyl ether	6.216	45	209864	18.299	ug/l #	88
23) Vinyl Acetate	6.157	43	525562	106.220	ug/l #	81
24) 1,1-Dichloroethane	6.110	63	113398	19.055	ug/l	93
25) 2-Butanone	7.087	43	105601	94.508	ug/l #	76
26) 2,2-Dichloropropane	7.075	77	99115	18.794	ug/l	90
27) cis-1,2-Dichloroethene	7.081	96	60194	18.650	ug/l	69
28) Bromochloromethane	7.428	49	41050	16.680	ug/l #	56
29) Tetrahydrofuran	7.452	42	61603	92.957	ug/l #	70
30) Chloroform	7.593	83	107149	19.284	ug/l	96
31) Cyclohexane	7.881	56	84933	17.162	ug/l #	80
32) 1,1,1-Trichloroethane	7.799	97	88682	19.371	ug/l	96
36) 1,1-Dichloropropene	8.010	75	72753	17.867	ug/l	92
37) Ethyl Acetate	7.163	43	42800	18.315	ug/l #	91
38) Carbon Tetrachloride	7.999	117	74830	19.501	ug/l	94
39) Methylcyclohexane	9.275	83	79540	17.120	ug/l #	78
40) Benzene	8.251	78	221422	19.069	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110623\
 Data File : VD077507.D
 Acq On : 06 Nov 2023 12:01
 Operator : JC/SY
 Sample : VD1106SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Nov 07 02:08:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	28754	19.230	ug/l #	71
42) 1,2-Dichloroethane	8.328	62	67140	18.256	ug/l	92
43) Isopropyl Acetate	8.357	43	87317	18.106	ug/l #	80
44) Trichloroethene	9.028	130	52198	18.352	ug/l	83
45) 1,2-Dichloropropane	9.310	63	66949	19.585	ug/l	99
46) Dibromomethane	9.393	93	29326	17.826	ug/l	91
47) Bromodichloromethane	9.587	83	80526	18.519	ug/l	99
48) Methyl methacrylate	9.381	41	37902	16.465	ug/l #	66
49) 1,4-Dioxane	9.387	88	6939	392.539	ug/l #	55
51) 4-Methyl-2-Pentanone	10.157	43	223087	93.984	ug/l #	80
52) Toluene	10.334	92	135104	19.086	ug/l	99
53) t-1,3-Dichloropropene	10.551	75	84321	18.313	ug/l	96
54) cis-1,3-Dichloropropene	10.016	75	98894	18.665	ug/l #	79
55) 1,1,2-Trichloroethane	10.734	97	42777	19.490	ug/l	94
56) Ethyl methacrylate	10.598	69	59548	18.247	ug/l #	64
57) 1,3-Dichloropropane	10.881	76	78208	19.064	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.869	63	143910	88.086	ug/l #	84
59) 2-Hexanone	10.922	43	169386	96.869	ug/l	81
60) Dibromochloromethane	11.075	129	50091	18.979	ug/l	96
61) 1,2-Dibromoethane	11.181	107	38200	18.378	ug/l	96
64) Tetrachloroethene	10.810	164	41047	18.424	ug/l #	89
65) Chlorobenzene	11.610	112	142220	19.400	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.681	131	53337	19.253	ug/l	97
67) Ethyl Benzene	11.687	91	278489	19.743	ug/l	93
68) m/p-Xylenes	11.792	106	199797	39.225	ug/l	80
69) o-Xylene	12.122	106	94616	19.393	ug/l	79
70) Styrene	12.134	104	165183	19.625	ug/l #	89
71) Bromoform	12.298	173	28529	18.575	ug/l #	100
73) Isopropylbenzene	12.422	105	263226	20.088	ug/l	94
74) N-amyl acetate	12.234	43	84412	18.187	ug/l #	81
75) 1,1,2,2-Tetrachloroethane	12.675	83	51422	20.019	ug/l	99
76) 1,2,3-Trichloropropane	12.728	75	66645	34.070	ug/l #	100
77) Bromobenzene	12.698	156	55854	19.365	ug/l	73
78) n-propylbenzene	12.763	91	332577	19.988	ug/l	93
79) 2-Chlorotoluene	12.851	91	180906	19.203	ug/l	91
80) 1,3,5-Trimethylbenzene	12.904	105	217481	19.780	ug/l	91
81) trans-1,4-Dichloro-2-b...	12.469	75	18982	18.744	ug/l	87
82) 4-Chlorotoluene	12.945	91	189345	19.295	ug/l	88
83) tert-Butylbenzene	13.169	119	189035	20.449	ug/l	89
84) 1,2,4-Trimethylbenzene	13.210	105	216260	19.800	ug/l	91
85) sec-Butylbenzene	13.345	105	295654	20.903	ug/l	94
86) p-Isopropyltoluene	13.463	119	232092	20.187	ug/l	93
87) 1,3-Dichlorobenzene	13.457	146	113601	20.037	ug/l	96
88) 1,4-Dichlorobenzene	13.539	146	112058	19.782	ug/l	96
89) n-Butylbenzene	13.786	91	243215	20.353	ug/l	97
90) Hexachloroethane	14.057	117	46728	19.862	ug/l	94
91) 1,2-Dichlorobenzene	13.834	146	99797	20.029	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.451	75	7950	18.372	ug/l	79
93) 1,2,4-Trichlorobenzene	15.098	180	63423	18.529	ug/l	98
94) Hexachlorobutadiene	15.204	225	35284	19.110	ug/l	97
95) Naphthalene	15.333	128	123839	18.965	ug/l	97
96) 1,2,3-Trichlorobenzene	15.522	180	56542	19.137	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110623\
Data File : VD077507.D
Acq On : 06 Nov 2023 12:01
Operator : JC/SY
Sample : VD1106SBSD01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :


Quant Time: Nov 07 02:08:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 02:08:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1069
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:168 Resp: 223882
Ion Ratio Lower Upper
168 100
99 57.6 36.7 55.1#



#2
Dichlorodifluoromethane
Concen: 19.315 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 85 Resp: 34868
Ion Ratio Lower Upper
85 100
87 32.3 15.7 47.0









#3
Chloromethane
Concen: 18.496 ug/l
RT: 2.152 min Scan# 90
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 50 Resp: 53011
Ion Ratio Lower Upper
50 100
52 32.7 26.6 40.0



#4
Vinyl Chloride
Concen: 22.207 ug/l
RT: 2.287 min Scan# 119
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 62 Resp: 62065
Ion Ratio Lower Upper
62 100
64 32.0 26.0 39.0









#5
Bromomethane
Concen: 23.919 ug/l
RT: 2.687 min Scan# 11
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 94 Resp: 35029
Ion Ratio Lower Upper
94 100
96 92.3 75.8 113.8



#6
Chloroethane
Concen: 23.451 ug/l
RT: 2.840 min Scan# 213
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Tgt Ion: 64 Resp: 44126
Ion Ratio Lower Upper
64 100
66 32.8 25.5 38.3








#7
 Trichlorofluoromethane
 Concen: 19.039 ug/l
 RT: 3.175 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Instrument :
 MSVOA_D
 ClientSampleId :







Tgt Ion:101 Resp: 74517
 Ion Ratio Lower Upper
 101 100
 103 63.1 52.3 78.5



#8
 Diethyl Ether
 Concen: 18.258 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Tgt Ion: 74 Resp: 25687
 Ion Ratio Lower Upper
 74 100
 45 113.5 37.5 112.7#









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.905 ug/l
RT: 3.969 min Scan# 405
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 47031
Ion Ratio Lower Upper
101 100
85 44.8 33.8 50.6
151 68.1 59.8 89.6



#10
Methyl Iodide
Concen: 17.167 ug/l
RT: 4.163 min Scan# 438
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:142 Resp: 44644
Ion Ratio Lower Upper
142 100
127 42.9 29.4 44.0
141 14.4 11.1 16.7









#11
Tert butyl alcohol
Concen: 92.566 ug/l
RT: 5.052 min Scan# 51
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 59 Resp: 23213
Ion Ratio Lower Upper
59 100
57 10.7 7.8 11.8



#12
1,1-Dichloroethene
Concen: 18.095 ug/l
RT: 3.940 min Scan# 400
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 96 Resp: 40477
Ion Ratio Lower Upper
96 100
61 172.5 100.6 150.8#
98 61.6 50.2 75.2





#13

Acrolein

Concen: 92.635 ug/l

RT: 3.799 min Scan# 31

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 56 Resp: 30634

Ion Ratio Lower Upper

56 100

55 71.7 55.1 82.7



#14

Allyl chloride

Concen: 17.853 ug/l

RT: 4.558 min Scan# 505

Delta R.T. -0.006 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion: 41 Resp: 96337


Ion Ratio Lower Upper

41 100

39 67.7 53.8 80.6

76 31.7 45.8 68.8#









#15
Acrylonitrile
Concen: 98.683 ug/l
RT: 5.258 min Scan# 61
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 53 Resp: 75051
Ion Ratio Lower Upper
53 100
52 82.5 65.0 97.4
51 36.4 28.3 42.5



#16
Acetone
Concen: 103.228 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 43 Resp: 86475
Ion Ratio Lower Upper
43 100
58 31.3 34.7 52.1#









#17
Carbon Disulfide
Concen: 14.448 ug/l
RT: 4.275 min Scan# 457
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 76 Resp: 79760
Ion Ratio Lower Upper
76 100
78 10.1 6.8 10.2



#18
Methyl Acetate
Concen: 17.952 ug/l
RT: 4.563 min Scan# 506
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 43 Resp: 54854
Ion Ratio Lower Upper
43 100
74 21.3 26.4 39.6#









#19
Methyl tert-butyl Ether
Concen: 18.821 ug/l
RT: 5.316 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 73 Resp: 132150
Ion Ratio Lower Upper
73 100
57 22.5 16.6 25.0



#20
Methylene Chloride
Concen: 21.974 ug/l
RT: 4.805 min Scan# 547
Delta R.T. -0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 84 Resp: 73469
Ion Ratio Lower Upper
84 100
49 152.0 71.4 107.2#
51 44.2 24.5 36.7#
86 60.1 48.7 73.1









#21
trans-1,2-Dichloroethene
Concen: 17.731 ug/l
RT: 5.316 min Scan# 61
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 96 Resp: 46192
Ion Ratio Lower Upper
96 100
61 142.8 89.3 133.9#
98 68.2 48.6 73.0



#22
Diisopropyl ether
Concen: 18.299 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 45 Resp: 209864
Ion Ratio Lower Upper
45 100
43 53.3 40.0 60.0
87 21.7 28.2 42.4#
59 11.0 11.0 16.4









#23
 Vinyl Acetate
 Concen: 106.220 ug/l
 RT: 6.157 min Scan# 777
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 43 Resp: 525562
 Ion Ratio Lower Upper
 43 100
 86 7.7 12.5 18.7#



#24
 1,1-Dichloroethane
 Concen: 19.055 ug/l
 RT: 6.110 min Scan# 769
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Tgt Ion: 63 Resp: 113398
 Ion Ratio Lower Upper
 63 100
 98 5.8 4.5 13.7
 100 4.2 2.5 7.5









#25
2-Butanone
Concen: 94.508 ug/l
RT: 7.087 min Scan# 91
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 105601
Ion Ratio Lower Upper
43 100
72 23.9 30.7 46.1#



#26
2,2-Dichloropropane
Concen: 18.794 ug/l
RT: 7.075 min Scan# 933
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Tgt Ion: 77 Resp: 99115
Ion Ratio Lower Upper
77 100
97 21.1 13.2 39.6








#27
 cis-1,2-Dichloroethene
 Concen: 18.650 ug/l
 RT: 7.081 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Instrument :
 MSVOA_D
 ClientSampleId :







Tgt Ion: 96 Resp: 60194
 Ion Ratio Lower Upper
 96 100
 61 171.1 0.0 239.2
 98 62.8 0.0 128.6



#28
 Bromochloromethane
 Concen: 16.680 ug/l
 RT: 7.428 min Scan# 993
 Delta R.T. 0.006 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Tgt Ion: 49 Resp: 41050
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 6.0
 130 57.5 83.2 124.8#





#29

Tetrahydrofuran

Concen: 92.957 ug/l

RT: 7.452 min Scan# 996

Delta R.T. 0.006 min


Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Instrument :

MSVOA_D

ClientSampleId :



Tgt Ion: 42	Resp: 61603
Ion Ratio	Lower Upper
42	100
72	35.1 46.5 69.7#
71	34.9 44.5 66.7#



#30

Chloroform

Concen: 19.284 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. -0.006 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Tgt Ion: 83	Resp: 107149
Ion Ratio	Lower Upper
83	100
85	63.6 53.4 80.2





#31

Cyclohexane

Concen: 17.162 ug/l

RT: 7.881 min Scan# 1070

Delta R.T. 0.006 min


Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :






Tgt Ion: 56 Resp: 84933

Ion Ratio Lower Upper

56 100

69 38.2 28.6 43.0

84 79.8 84.8 127.2#



#32

1,1,1-Trichloroethane

Concen: 19.371 ug/l

RT: 7.799 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion: 97 Resp: 88682


Ion Ratio Lower Upper

97 100

99 63.9 52.0 78.0

61 48.0 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 45.959 ug/l

RT: 8.234 min Scan# 1130

Delta R.T. 0.006 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 65 Resp: 145496

Ion Ratio Lower Upper

65 100

67 54.7 0.0 115.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion: 114 Resp: 397362


Ion Ratio Lower Upper

114 100

63 24.4 0.0 33.4


88 15.5 0.0 26.2








#35
Dibromofluoromethane
Concen: 46.754 ug/l
RT: 7.804 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:113 Resp: 128407
Ion Ratio Lower Upper
113 100
111 102.7 84.7 127.1
192 19.2 16.8 25.2



#36
1,1-Dichloropropene
Concen: 17.867 ug/l
RT: 8.010 min Scan# 1092
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Tgt Ion: 75 Resp: 72753
Ion Ratio Lower Upper
75 100
110 32.8 19.9 59.6
77 32.7 24.7 37.1








#37
Ethyl Acetate
Concen: 18.315 ug/l
RT: 7.163 min Scan# 949
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 43 Resp: 42800
Ion Ratio Lower Upper
43 100
61 13.8 14.9 22.3#
70 9.5 10.0 15.0#



#38
Carbon Tetrachloride
Concen: 19.501 ug/l
RT: 7.999 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 117 Resp: 74830
Ion Ratio Lower Upper
117 100
119 90.1 78.0 117.0
121 29.1 24.2 36.2





#39

Methylcyclohexane

Concen: 17.120 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. 0.000 min


Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Instrument :

MSVOA_D

ClientSampleId :



Tgt Ion:	83	Resp:	79540
Ion Ratio	Lower	Upper	
83	100		
55	87.4	49.4	74.0#
98	43.1	38.4	57.6



#40

Benzene

Concen: 19.069 ug/l




RT: 8.251 min Scan# 1133


Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01


Tgt Ion:	78	Resp:	221422
Ion Ratio	Lower	Upper	
78	100		
77	26.2	19.4	29.0








#41
Methacrylonitrile
Concen: 19.230 ug/l
RT: 7.399 min Scan# 91
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:	41	Resp:	28754
Ion	Ratio	Lower	Upper
41	100		
39	55.7	34.0	51.0#
67	59.1	77.0	115.4#
52	27.0	27.9	41.9#



#42
1,2-Dichloroethane
Concen: 18.256 ug/l
RT: 8.328 min Scan# 1146
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Tgt Ion:	62	Resp:	67140
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	25.4








#43
Isopropyl Acetate
Concen: 18.106 ug/l
RT: 8.357 min Scan# 1151
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 43 Resp: 87317
Ion Ratio Lower Upper
43 100
61 26.6 31.4 47.0#
87 10.8 14.9 22.3#



#44
Trichloroethene
Concen: 18.352 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:130 Resp: 52198
Ion Ratio Lower Upper
130 100
95 105.2 0.0 179.2





#45

1,2-Dichloropropane

Concen: 19.585 ug/l

RT: 9.310 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 63 Resp: 66949

Ion Ratio Lower Upper

63 100

65 30.3 24.5 36.7



#46

Dibromomethane

Concen: 17.826 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion: 93 Resp: 29326


Ion Ratio Lower Upper

93 100

95 84.0 67.7 101.5

174 90.1 85.4 128.2









#47
Bromodichloromethane
Concen: 18.519 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
83	100		
85	63.9	50.2	75.4
127	7.9	6.3	9.5



#48
Methyl methacrylate
Concen: 16.465 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion	Ratio	Lower	Upper
41	100		
69	78.1	102.2	153.2#
39	59.8	41.4	62.2





#49

1,4-Dioxane

Concen: 392.539 ug/l

RT: 9.387 min Scan# 11

Delta R.T. 0.006 min


Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :






Tgt Ion: 88 Resp: 6939

Ion Ratio Lower Upper

88 100

43 50.1 19.0 28.6#

58 83.7 43.5 65.3#



#50

Toluene-d8

Concen: 49.995 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VD077507.D




Acq: 06 Nov 2023 12:01


Tgt Ion: 98 Resp: 533339

Ion Ratio Lower Upper

98 100

100 63.4 50.3 75.5









#51
4-Methyl-2-Pentanone
Concen: 93.984 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 223087
Ion Ratio Lower Upper
43 100
58 35.3 38.9 58.3#



#52
Toluene
Concen: 19.086 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 92 Resp: 135104
Ion Ratio Lower Upper
92 100
91 177.1 143.0 214.4









#53
t-1,3-Dichloropropene
Concen: 18.313 ug/l
RT: 10.551 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 84321
Ion Ratio Lower Upper
75 100
77 33.8 25.3 37.9



#54
cis-1,3-Dichloropropene
Concen: 18.665 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 75 Resp: 98894
Ion Ratio Lower Upper
75 100
77 30.0 25.5 38.3
39 49.0 22.1 33.1#









#55
1,1,2-Trichloroethane
Concen: 19.490 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:	97	Resp:	42777
Ion	Ratio	Lower	Upper
97	100		
83	89.5	65.8	98.6
85	58.0	45.7	68.5
99	63.6	46.2	69.4



#56
Ethyl methacrylate
Concen: 18.247 ug/l
RT: 10.598 min Scan# 1532
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Tgt Ion:	69	Resp:	59548
Ion	Ratio	Lower	Upper
69	100		
41	77.4	39.1	58.7#
39	36.8	20.0	30.0#








#57
1,3-Dichloropropane
Concen: 19.064 ug/l
RT: 10.881 min Scan# 1580
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 76 Resp: 78208
Ion Ratio Lower Upper
76 100
78 32.3 26.4 39.6



#58
2-Chloroethyl Vinyl ether
Concen: 88.086 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 63 Resp: 143910
Ion Ratio Lower Upper
63 100
106 22.7 25.3 37.9#









#59
2-Hexanone
Concen: 96.869 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 169386
Ion Ratio Lower Upper
43 100
58 52.4 34.1 102.2



#60
Dibromochloromethane
Concen: 18.979 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 129 Resp: 50091
Ion Ratio Lower Upper
129 100
127 77.8 37.1 111.5









#61
1,2-Dibromoethane
Concen: 18.378 ug/l
RT: 11.181 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:107 Resp: 38200
Ion Ratio Lower Upper
107 100
109 90.7 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 46.342 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 95 Resp: 166505
Ion Ratio Lower Upper
95 100
174 80.3 0.0 201.8
176 77.2 0.0 191.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:117 Resp: 349930

Ion Ratio Lower Upper

117 100

82 61.4 38.3 57.5#

119 32.5 25.0 37.4



#64

Tetrachloroethene

Concen: 18.424 ug/l

RT: 10.810 min Scan# 1568

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Tgt Ion:164 Resp: 41047




Ion Ratio Lower Upper


164 100

166 130.5 99.2 148.8

129 95.4 68.4 102.6

131 97.0 63.8 95.8#





#65

Chlorobenzene

Concen: 19.400 ug/l

RT: 11.610 min Scan# 1704

Delta R.T. 0.006 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion:112 Resp: 142220

Ion Ratio Lower Upper

112 100

114 33.3 25.1 37.7



#66

1,1,1,2-Tetrachloroethane

Concen: 19.253 ug/l

RT: 11.681 min Scan# 1716

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion:131 Resp: 53337


Ion Ratio Lower Upper

131 100

133 94.5 48.5 145.5

119 68.4 35.6 106.9









#67
Ethyl Benzene
Concen: 19.743 ug/l
RT: 11.687 min Scan# 1716
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 278489
Ion Ratio Lower Upper
91 100
106 29.4 26.5 39.7



#68
m/p-Xylenes
Concen: 39.225 ug/l
RT: 11.792 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:106 Resp: 199797
Ion Ratio Lower Upper
106 100
91 212.1 147.0 220.6









#69
o-Xylene
Concen: 19.393 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:	106	Resp:	94616
Ion Ratio	Lower	Upper	
106	100		
91	227.7	97.9	293.6



#70
Styrene
Concen: 19.625 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:	104	Resp:	165183
Ion Ratio	Lower	Upper	
104	100		
78	53.6	33.2	49.8#
103	57.4	43.6	65.4





#71

Bromoform

Concen: 18.575 ug/l

RT: 12.298 min Scan# 1821

Delta R.T. 0.000 min

Lab File: VD077507.D





Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :

Tgt Ion:173	Resp:	28529	
Ion Ratio	Lower	Upper	
173	100		
175	49.2	24.6	73.8
254	0.0	0.0	0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.516 min Scan# 2028


Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01


Tgt Ion:152	Resp:	160722	
Ion Ratio	Lower	Upper	
152	100		
115	60.7	27.1	81.3
150	160.9	0.0	353.6








#73
Isopropylbenzene
Concen: 20.088 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion:105 Resp: 263226
Ion Ratio Lower Upper
105 100
120 25.6 14.3 42.9




#74
N-ethyl acetate
Concen: 18.187 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01



Tgt Ion: 43 Resp: 84412
Ion Ratio Lower Upper
43 100
70 37.1 45.0 67.4#
55 22.1 23.8 35.6#
61 23.3 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 20.019 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 51422
Ion Ratio Lower Upper
83 100
131 8.5 5.1 15.3
85 65.3 32.6 97.8



#76
1,2,3-Trichloropropane
Concen: 34.070 ug/l
RT: 12.728 min Scan# 1894
Delta R.T. 0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 75 Resp: 66645
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 19.365 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. -0.006 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:156 Resp: 55854
Ion Ratio Lower Upper
156 100
77 210.2 76.8 230.4
158 96.5 48.6 145.9



#78
n-propylbenzene
Concen: 19.988 ug/l
RT: 12.763 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 91 Resp: 332577
Ion Ratio Lower Upper
91 100
120 21.5 12.4 37.2









#79
2-Chlorotoluene
Concen: 19.203 ug/l
RT: 12.851 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 180906
Ion Ratio Lower Upper
91 100
126 33.0 19.4 58.1



#80
1,3,5-Trimethylbenzene
Concen: 19.780 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 105 Resp: 217481
Ion Ratio Lower Upper
105 100
120 46.3 26.4 79.2









#81
trans-1,4-Dichloro-2-butene
Concen: 18.744 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 18982
Ion Ratio Lower Upper
75 100
53 93.5 63.0 94.4
89 42.2 37.9 56.9



#82
4-Chlorotoluene
Concen: 19.295 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion: 91 Resp: 189345
Ion Ratio Lower Upper
91 100
126 31.7 19.4 58.4





#83

tert-Butylbenzene

Concen: 20.449 ug/l

RT: 13.169 min Scan# 1969

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:119 Resp: 189035

Ion Ratio Lower Upper

119 100

91 69.0 29.3 87.8

134 25.2 14.1 42.3



#84

1,2,4-Trimethylbenzene

Concen: 19.800 ug/l

RT: 13.210 min Scan# 1976

Delta R.T. 0.000 min

Lab File: VD077507.D




Acq: 06 Nov 2023 12:01


Tgt Ion:105 Resp: 216260

Ion Ratio Lower Upper

105 100

120 42.3 24.3 72.9









#85
 sec-Butylbenzene
 Concen: 20.903 ug/l
 RT: 13.345 min Scan# 1999
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:105 Resp: 295654
 Ion Ratio Lower Upper
 105 100
 134 19.2 10.9 32.9



#86
 p-Isopropyltoluene
 Concen: 20.187 ug/l
 RT: 13.463 min Scan# 2019
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01


Tgt Ion:119 Resp: 232092
 Ion Ratio Lower Upper
 119 100
 134 25.2 14.2 42.6
 91 24.6 10.5 31.5








#87
1,3-Dichlorobenzene
Concen: 20.037 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:146 Resp: 113601
Ion Ratio Lower Upper
146 100
111 40.3 17.9 53.8
148 63.4 32.3 96.8



#88
1,4-Dichlorobenzene
Concen: 19.782 ug/l
RT: 13.539 min Scan# 2032
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:146 Resp: 112058
Ion Ratio Lower Upper
146 100
111 40.5 17.4 52.3
148 64.3 31.7 95.1





#89

n-Butylbenzene

Concen: 20.353 ug/l

RT: 13.786 min Scan# 2074

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion: 91 Resp: 243215

Ion Ratio Lower Upper

91 100

92 52.6 26.5 79.5

134 23.3 13.8 41.3



#90

Hexachloroethane

Concen: 19.862 ug/l

RT: 14.057 min Scan# 2120

Delta R.T. 0.006 min

Lab File: VD077507.D




Acq: 06 Nov 2023 12:01


Tgt Ion: 117 Resp: 46728

Ion Ratio Lower Upper

117 100

201 71.4 38.5 115.5









#91
 1,2-Dichlorobenzene
 Concen: 20.029 ug/l
 RT: 13.834 min Scan# 2082
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:146 Resp: 99797
 Ion Ratio Lower Upper
 146 100
 111 40.8 18.3 54.8
 148 63.1 31.6 94.7



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 18.372 ug/l
 RT: 14.451 min Scan# 2187
 Delta R.T. 0.000 min
 Lab File: VD077507.D
 Acq: 06 Nov 2023 12:01

Tgt Ion: 75 Resp: 7950
 Ion Ratio Lower Upper
 75 100
 155 76.3 47.9 143.8
 157 96.1 59.9 179.7





#93

1,2,4-Trichlorobenzene

Concen: 18.529 ug/l

RT: 15.098 min Scan# 2297

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:180 Resp: 63423

Ion Ratio Lower Upper

180 100

182 98.1 48.5 145.6

145 30.8 14.2 42.6



#94

Hexachlorobutadiene

Concen: 19.110 ug/l

RT: 15.204 min Scan# 2315

Delta R.T. 0.000 min

Lab File: VD077507.D

Acq: 06 Nov 2023 12:01




Tgt Ion:225 Resp: 35284


Ion Ratio Lower Upper

225 100

223 64.7 31.6 94.8

227 66.5 31.7 95.1









#95
Naphthalene
Concen: 18.965 ug/l
RT: 15.333 min Scan# 21
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:	128	Resp:	123839
Ion Ratio	Lower	Upper	
128	100		
127	11.7	10.5	15.7
129	10.0	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.137 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. 0.000 min
Lab File: VD077507.D
Acq: 06 Nov 2023 12:01

Tgt Ion:	180	Resp:	56542
Ion Ratio	Lower	Upper	
180	100		
182	94.6	47.3	141.9
145	30.6	15.1	45.3

