

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110623\
 Data File : VD077508.D
 Acq On : 06 Nov 2023 14:43
 Operator : JC/SY
 Sample : 05256-03
 Misc : 5.46G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1-VOC

Quant Time: Nov 07 02:09:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 04:19:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	274867	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	525976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	470646	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	209699	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	167609	43.123	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.240%
35) Dibromofluoromethane	7.804	113	158178	43.511	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.020%
50) Toluene-d8	10.269	98	679776	48.141	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.280%
62) 4-Bromofluorobenzene	12.575	95	209960	44.147	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110623\
Data File : VD077508.D
Acq On : 06 Nov 2023 14:43
Operator : JC/SY
Sample : 05256-03
Misc : 5.46G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1-VOC

Quant Time: Nov 07 02:09:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D103023S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 04:19:17 2023
Response via : Initial Calibration

