

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110821\
 Data File : VD070936.D
 Acq On : 08 Nov 2021 15:26
 Operator : VA/SY
 Sample : M4519-01
 Misc : 5.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Nov 09 02:33:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 08 13:56:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	495468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	924938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	852975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	390464	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	276948	45.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.020%
35) Dibromofluoromethane	7.914	113	276178	47.663	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.320%
50) Toluene-d8	10.332	98	1147160	50.930	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.860%
62) 4-Bromofluorobenzene	12.620	95	424645	49.940	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110821\
Data File : VD070936.D
Acq On : 08 Nov 2021 15:26
Operator : VA/SY
Sample : M4519-01
Misc : 5.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Nov 09 02:33:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110521S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 08 13:56:36 2021
Response via : Initial Calibration

