

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110821\
 Data File : VD070932.D
 Acq On : 08 Nov 2021 13:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 02:32:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 08 13:56:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.537	0.503	6.3	98	0.00
3 P	Chloromethane	0.532	0.510	4.1	102	0.00
4 C	Vinyl Chloride	0.515	0.522	-1.4#	103	0.00
5 T	Bromomethane	0.314	0.328	-4.5	110	0.00
6 T	Chloroethane	0.322	0.330	-2.5	105	0.00
7 T	Trichlorofluoromethane	0.947	0.973	-2.7	101	0.00
8 T	Diethyl Ether	0.267	0.303	-13.5	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.541	0.587	-8.5	106	0.00
10 T	Methyl Iodide	0.529	0.609	-15.1	100	0.00
11 T	Tert butyl alcohol	0.040	0.041	-2.5	104	0.00
12 CM	1,1-Dichloroethene	0.522	0.573	-9.8#	109	0.00
13 T	Acrolein	0.012	0.014	-16.7	116	0.00
14 T	Allyl chloride	1.097	1.213	-10.6	107	0.00
15 T	Acrylonitrile	0.119	0.142	-19.3	109	0.00
16 T	Acetone	0.144	0.182	-26.4#	116	0.00
17 T	Carbon Disulfide	1.683	1.830	-8.7	106	0.00
18 T	Methyl Acetate	0.451	0.523	-16.0	109	0.00
19 T	Methyl tert-butyl Ether	1.298	1.441	-11.0	106	0.00
20 T	Methylene Chloride	0.765	0.682	10.8	112	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.634	-7.3	105	0.00
22 T	Diisopropyl ether	2.005	2.229	-11.2	108	0.00
23 T	Vinyl Acetate	0.940	1.122	-19.4	111	0.00
24 P	1,1-Dichloroethane	1.141	1.234	-8.2	106	0.00
25 T	2-Butanone	0.170	0.207	-21.8	111	0.00
26 T	2,2-Dichloropropane	1.052	1.127	-7.1	105	0.00
27 T	cis-1,2-Dichloroethene	0.660	0.719	-8.9	107	0.00
28 T	Bromochloromethane	0.456	0.498	-9.2	110	0.00
29 T	Tetrahydrofuran	0.104	0.122	-17.3	109	0.00
30 C	Chloroform	1.106	1.185	-7.1#	105	0.00
31 T	Cyclohexane	1.201	1.250	-4.1	107	0.00
32 T	1,1,1-Trichloroethane	1.022	1.054	-3.1	102	0.00
33 S	1,2-Dichloroethane-d4	0.614	0.586	4.6	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.313	0.303	3.2	104	0.00
36 T	1,1-Dichloropropene	0.511	0.543	-6.3	105	0.00
37 T	Ethyl Acetate	0.209	0.246	-17.7	108	0.00
38 T	Carbon Tetrachloride	0.496	0.506	-2.0	100	0.00
39 T	Methylcyclohexane	0.656	0.690	-5.2	105	0.00
40 TM	Benzene	1.355	1.470	-8.5	108	0.00
41 T	Methacrylonitrile	0.137	0.165	-20.4	125	0.00
42 TM	1,2-Dichloroethane	0.417	0.440	-5.5	102	0.00
43 T	Isopropyl Acetate	0.447	0.514	-15.0	111	0.00
44 TM	Trichloroethene	0.364	0.383	-5.2	104	0.00
45 C	1,2-Dichloropropane	0.356	0.393	-10.4#	108	0.00
46 T	Dibromomethane	0.176	0.189	-7.4	105	0.00
47 T	Bromodichloromethane	0.481	0.513	-6.7	105	0.00
48 T	Methyl methacrylate	0.231	0.260	-12.6	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110821\
 Data File : VD070932.D
 Acq On : 08 Nov 2021 13:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 02:32:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 08 13:56:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	102	0.00
50 S	Toluene-d8	1.218	1.158	4.9	102	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.246	-17.1	108	0.00
52 CM	Toluene	0.866	0.918	-6.0#	105	0.00
53 T	t-1,3-Dichloropropene	0.479	0.515	-7.5	106	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.599	-7.3	106	0.00
55 T	1,1,2-Trichloroethane	0.230	0.253	-10.0	107	0.00
56 T	Ethyl methacrylate	0.337	0.375	-11.3	107	0.00
57 T	1,3-Dichloropropane	0.425	0.474	-11.5	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.161	0.180	-11.8	114	0.00
59 T	2-Hexanone	0.156	0.183	-17.3	110	0.00
60 T	Dibromochloromethane	0.294	0.315	-7.1	104	0.00
61 T	1,2-Dibromoethane	0.229	0.249	-8.7	105	0.00
62 S	4-Bromofluorobenzene	0.460	0.421	8.5	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.333	0.346	-3.9	104	0.00
65 PM	Chlorobenzene	0.986	1.052	-6.7	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.381	-5.5	105	0.00
67 C	Ethyl Benzene	1.906	2.010	-5.5#	105	0.00
68 T	m/p-Xylenes	0.705	0.742	-5.2	105	0.00
69 T	o-Xylene	0.675	0.708	-4.9	104	0.00
70 T	Styrene	1.116	1.193	-6.9	105	0.00
71 P	Bromoform	0.179	0.194	-8.4	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.106	4.355	-6.1	103	0.00
74 T	N-amyl acetate	1.077	1.248	-15.9	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.627	0.728	-16.1	108	0.00
76 T	1,2,3-Trichloropropane	0.504	0.476	5.6	87	0.00
77 T	Bromobenzene	0.847	0.914	-7.9	105	0.00
78 T	n-propylbenzene	5.015	5.419	-8.1	105	0.00
79 T	2-Chlorotoluene	2.880	3.067	-6.5	105	0.00
80 T	1,3,5-Trimethylbenzene	3.406	3.556	-4.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.212	0.248	-17.0	106	0.00
82 T	4-Chlorotoluene	2.976	3.128	-5.1	104	0.00
83 T	tert-Butylbenzene	2.935	3.049	-3.9	102	0.00
84 T	1,2,4-Trimethylbenzene	3.312	3.486	-5.3	104	0.00
85 T	sec-Butylbenzene	4.356	4.617	-6.0	105	0.00
86 T	p-Isopropyltoluene	3.564	3.721	-4.4	103	0.00
87 T	1,3-Dichlorobenzene	1.656	1.775	-7.2	105	0.00
88 T	1,4-Dichlorobenzene	1.641	1.762	-7.4	104	0.00
89 T	n-Butylbenzene	3.520	3.748	-6.5	105	0.00
90 T	Hexachloroethane	0.701	0.745	-6.3	105	0.00
91 T	1,2-Dichlorobenzene	1.423	1.546	-8.6	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.116	0.124	-6.9	101	0.00
93 T	1,2,4-Trichlorobenzene	0.918	0.993	-8.2	104	0.00
94 T	Hexachlorobutadiene	0.497	0.521	-4.8	102	0.00
95 T	Naphthalene	1.704	1.945	-14.1	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110821\
Data File : VD070932.D
Acq On : 08 Nov 2021 13:23
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 09 02:32:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110521S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 08 13:56:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.770	0.846	-9.9	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6