

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD110918\
 Data File : VD060327.D
 Acq On : 9 Nov 2018 22:15
 Operator : VA/AP
 Sample : J5892-01
 Misc : 5.06g/5ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PN-STONE

Quant Time: Nov 12 01:37:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	14519	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.45	114	24985	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.29	117	26903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.38	152	17172	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.79	65	19933	191.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	383.48%	
35) Dibromofluoromethane	6.34	113	13960	74.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	149.86%	
50) Toluene-d8	9.47	98	25381	46.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.10%	
62) 4-Bromofluorobenzene	12.43	95	13726	69.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	138.94%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.16	94	568	21.316	ug/l #	2
13) Acrolein	2.91	56	178	35.288	ug/l #	14
14) Allyl chloride	3.54	41	1366	10.077	ug/l #	86
15) Acrylonitrile	4.11	53	4851	194.867	ug/l	89
16) Acetone	3.17	43	5599	310.492	ug/l #	55
18) Methyl Acetate	3.57	43	1957	50.897	ug/l #	63
20) Methylene Chloride	3.73	84	2263	14.147	ug/l #	77
25) 2-Butanone	5.66	43	1027	27.400	ug/l #	60
27) cis-1,2-Dichloroethene	5.62	96	437	2.667	ug/l	1
29) Tetrahydrofuran	5.96	42	1197	60.572	ug/l #	41
36) 1,1-Dichloropropene	6.42	75	688	2.787	ug/l #	41
37) Ethyl Acetate	5.66	43	1027	9.042	ug/l #	69
38) Carbon Tetrachloride	6.41	117	901	3.935	ug/l #	12
41) Methacrylonitrile	5.97	41	199	1.531	ug/l #	42
42) 1,2-Dichloroethane	6.91	62	212	1.285	ug/l	71
44) Trichloroethene	7.75	130	988	5.173	ug/l	1
46) Dibromomethane	8.23	93	701	7.532	ug/l	90
53) t-1,3-Dichloropropene	9.92	75	459	2.477	ug/l #	13
55) 1,1,2-Trichloroethane	10.19	97	278	2.383	ug/l #	15
58) 2-Chloroethyl Vinyl ether	8.94	63	766	11.811	ug/l #	48
61) 1,2-Dibromoethane	10.77	107	396	3.388	ug/l #	58
76) 1,2,3-Trichloropropane	12.43	75	4692	22.841	ug/l #	42
81) trans-1,4-Dichloro-2-buten	12.43	75	5852	100.065	ug/l #	15
89) n-Butylbenzene	13.65	91	551	Below Cal	#	29
95) Naphthalene	14.97	128	994	2.114	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD110918\
Data File : VD060327.D
Acq On : 9 Nov 2018 22:15
Operator : VA/AP
Sample : J5892-01
Misc : 5.06g/5ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PN-STONE

Quant Time: Nov 12 01:37:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 6.42 min Scan# 497

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :


PN-STONE

Tgt Ion:168 Resp: 14519

Ion Ratio Lower Upper




168 100

99 47.1 39.8 59.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#5

Bromomethane

Concen: 21.316 ug/l

RT: 2.16 min Scan# 64

Delta R.T. -0.01 min

Lab File: VD060327.D


Acq: 9 Nov 2018 22:15

Tgt Ion: 94 Resp: 568

Ion Ratio Lower Upper



94 100


96 0.0 76.6 114.8#



Abundance Ion 93.90 (93.60 to 94.60): VDC

Ion 95.90 (95.60 to 96.60): VDC





#13

Acrolein

Concen: 35.288 ug/l

RT: 2.91 min Scan# 141

Delta R.T. -0.08 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :


PN-STONE

Tgt Ion: 56 Resp: 178

Ion Ratio Lower Upper




56 100

55 0.0 57.0 85.6#



Abundance Ion 55.95 (55.65 to 56.65): VDC

Ion 55.00 (54.70 to 55.70): VDC



#14

Allyl chloride

Concen: 10.077 ug/l

RT: 3.54 min Scan# 205

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15


Tgt Ion: 41 Resp: 1366

Ion Ratio Lower Upper

41 100

39 75.3 60.4 90.6



76 0.0 21.8 32.8#




Abundance Ion 41.05 (40.75 to 41.75): VDC

Ion 38.95 (38.65 to 39.65): VDC

Ion 75.95 (75.65 to 76.65): VDC





#15

Acrylonitrile

Concen: 194.867 ug/l

RT: 4.11 min Scan# 262

Delta R.T. 0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :

PN-STONE


Tgt Ion: 53 Resp: 4851

Ion Ratio Lower Upper

53 100

52 68.9 67.0 100.4




51 37.6 29.9 44.9



Abundance Ion 53.00 (52.70 to 53.70): VDC

Ion 52.00 (51.70 to 52.70): VDC

Ion 51.00 (50.70 to 51.70): VDC



#16

Acetone

Concen: 310.492 ug/l

RT: 3.17 min Scan# 167

Delta R.T. 0.00 min

Lab File: VD060327.D


Acq: 9 Nov 2018 22:15

Tgt Ion: 43 Resp: 5599

Ion Ratio Lower Upper



43 100


58 0.0 16.9 25.3#



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC









#18
Methyl Acetate
Concen: 50.897 ug/l
RT: 3.57 min Scan# 208
Delta R.T. 0.01 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15




Instrument :
MSVOA_D
ClientSampleId :
PN-STONE


Tgt Ion: 43 Resp: 1957
Ion Ratio Lower Upper
43 100
74 0.0 12.4 18.6#



#20
Methylene Chloride
Concen: 14.147 ug/l
RT: 3.73 min Scan# 224
Delta R.T. 0.00 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Tgt Ion: 84 Resp: 2263
Ion Ratio Lower Upper
84 100
49 128.5 115.6 173.4
51 0.0 35.6 53.4#
86 53.2 52.3 78.5





#25

2-Butanone

Concen: 27.400 ug/l

RT: 5.66 min Scan# 420

Delta R.T. 0.02 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :


PN-STONE

Tgt Ion: 43 Resp: 1027

Ion Ratio Lower Upper




43 100

72 0.0 13.8 20.8#



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 72.00 (71.70 to 72.70): VDC



#27

cis-1,2-Dichloroethene

Concen: 2.667 ug/l

RT: 5.62 min Scan# 416

Delta R.T. 0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15


Tgt Ion: 96 Resp: 437

Ion Ratio Lower Upper

96 100

61 0.0 0.0 259.0



98 0.0 0.0 128.8




Abundance Ion 95.90 (95.60 to 96.60): VDC

Ion 60.95 (60.65 to 61.65): VDC

Ion 97.85 (97.55 to 98.55): VDC





#29

Tetrahydrofuran

Concen: 60.572 ug/l

RT: 5.96 min Scan# 450

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :

PN-STONE


Tgt Ion: 42 Resp: 1197

Ion Ratio Lower Upper

42 100




72 0.0 27.6 41.4#

71 0.0 26.5 39.7#



Abundance Ion 42.00 (41.70 to 42.70): VDC
Ion 72.00 (71.70 to 72.70): VDC
Ion 71.00 (70.70 to 71.70): VDC

5.96



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 6.79 min Scan# 535

Delta R.T. -0.01 min

Lab File: VD060327.D


Acq: 9 Nov 2018 22:15

Tgt Ion: 65 Resp: 19933

Ion Ratio Lower Upper



65 100


67 44.0 0.0 109.8



Abundance Ion 64.95 (64.65 to 65.65): VDC
Ion 66.90 (66.60 to 67.60): VDC

6.79









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 7.45 min Scan# 602
 Delta R.T. -0.01 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Instrument :
 MSVOA_D
 ClientSampled :
 PN-STONE

Tgt Ion:114 Resp: 24985
 Ion Ratio Lower Upper
 114 100
 63 20.3 0.0 34.0
 88 19.1 0.0 35.4




Abundance Ion 114.00 (113.70 to 114.70): V
 Ion 62.95 (62.65 to 63.65): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 6.34 min Scan# 489
 Delta R.T. 0.00 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Tgt Ion:113 Resp: 13960
 Ion Ratio Lower Upper
 113 100
 111 106.5 81.6 122.4
 192 21.4 16.5 24.7



Abundance Ion 112.90 (112.60 to 113.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 191.80 (191.50 to 192.50): V









#36
 1,1-Dichloropropene
 Concen: 2.787 ug/l
 RT: 6.42 min Scan# 497
 Delta R.T. -0.13 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Instrument :
 MSVOA_D
 ClientSampleId :
 PN-STONE

Tgt Ion: 75 Resp: 688
 Ion Ratio Lower Upper
 75 100
 110 0.0 18.0 54.0#
 77 0.0 25.6 38.4#




Abundance Ion 75.00 (74.70 to 75.70): VDC
 Ion 109.90 (109.60 to 110.60): VDC
 Ion 76.95 (76.65 to 77.65): VDC






#37
 Ethyl Acetate
 Concen: 9.042 ug/l
 RT: 5.66 min Scan# 420
 Delta R.T. -0.05 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Tgt Ion: 43 Resp: 1027
 Ion Ratio Lower Upper
 43 100
 61 0.0 11.8 17.6#
 70 0.0 6.2 9.4#



Abundance Ion 42.95 (42.65 to 43.65): VDC
 Ion 60.95 (60.65 to 61.65): VDC
 Ion 70.00 (69.70 to 70.70): VDC









#38
Carbon Tetrachloride
Concen: 3.935 ug/l
RT: 6.41 min Scan# 496
Delta R.T. -0.11 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Instrument :
MSVOA_D
ClientSampleId :
PN-STONE

Tgt Ion: 117 Resp: 901
Ion Ratio Lower Upper
117 100
119 0.0 76.6 115.0#
121 0.0 24.3 36.5#




Abundance Ion 116.85 (116.55 to 117.55): V
Ion 118.85 (118.55 to 119.55): V
Ion 120.85 (120.55 to 121.55): V






#41
Methacrylonitrile
Concen: 1.531 ug/l
RT: 5.97 min Scan# 451
Delta R.T. 0.00 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Tgt Ion: 41 Resp: 199
Ion Ratio Lower Upper
41 100
39 91.4 36.2 54.2#
67 0.0 11.9 17.9#
52 0.0 5.2 7.8#



Abundance Ion 40.95 (40.65 to 41.65): VDC
Ion 38.95 (38.65 to 39.65): VDC
Ion 67.00 (66.70 to 67.70): VDC
Ion 52.00 (51.70 to 52.70): VDC









#42
 1,2-Dichloroethane
 Concen: 1.285 ug/l
 RT: 6.91 min Scan# 547
 Delta R.T. 0.00 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15




Instrument :
 MSVOA_D
 ClientSampleId :
 PN-STONE


Tgt Ion: 62 Resp: 212
 Ion Ratio Lower Upper
 62 100
 98 0.0 0.0 22.0



#44
 Trichloroethene
 Concen: 5.173 ug/l
 RT: 7.75 min Scan# 632
 Delta R.T. -0.01 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Tgt Ion: 130 Resp: 988
 Ion Ratio Lower Upper
 130 100
 95 0.0 0.0 200.0









#46
Dibromomethane
Concen: 7.532 ug/l
RT: 8.23 min Scan# 681
Delta R.T. 0.00 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Instrument :
MSVOA_D
ClientSampleId :
PN-STONE

Tgt Ion: 93 Resp: 701
Ion Ratio Lower Upper
93 100
95 96.1 66.4 99.6
174 103.0 87.8 131.8




Abundance Ion 92.90 (92.60 to 93.60): VDC
Ion 94.90 (94.60 to 95.60): VDC
Ion 173.80 (173.50 to 174.50): VDC






#50
Toluene-d8
Concen: N.D. ug/l
RT: 9.47 min Scan# 807
Delta R.T. 0.00 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Tgt Ion: 98 Resp: 25381
Ion Ratio Lower Upper
98 100
100 74.3 55.2 82.8



Abundance Ion 98.05 (97.75 to 98.75): VDC
Ion 100.05 (99.75 to 100.75): VDC





#53

t-1,3-Dichloropropene

Concen: 2.477 ug/l

RT: 9.92 min Scan# 853

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :


PN-STONE

Tgt Ion: 75 Resp: 459

Ion Ratio Lower Upper




75 100

77 79.7 25.2 37.8#



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC



#55

1,1,2-Trichloroethane

Concen: 2.383 ug/l

RT: 10.19 min Scan# 880

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Tgt Ion: 97 Resp: 278


Ion Ratio Lower Upper

97 100

83 0.0 68.4 102.6#

85 0.0 46.8 70.2#

99 0.0 49.5 74.3#






Abundance Ion 96.85 (96.55 to 97.55): VDC

Ion 82.95 (82.65 to 83.65): VDC

Ion 84.95 (84.65 to 85.65): VDC

Ion 98.85 (98.55 to 99.55): VDC









#58
 2-Chloroethyl Vinyl ether
 Concen: 11.811 ug/l
 RT: 8.94 min Scan# 753
 Delta R.T. 0.00 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15




Instrument :
 MSVOA_D
 ClientSampleId :
 PN-STONE


Tgt Ion: 63 Resp: 766
 Ion Ratio Lower Upper
 63 100
 106 0.0 21.5 32.3#



#61
 1,2-Dibromoethane
 Concen: 3.388 ug/l
 RT: 10.77 min Scan# 939
 Delta R.T. 0.00 min
 Lab File: VD060327.D
 Acq: 9 Nov 2018 22:15

Tgt Ion: 107 Resp: 396
 Ion Ratio Lower Upper
 107 100
 109 135.4 75.7 113.5#









#62
4-Bromofluorobenzene
Concen: 3.388 ug/l
RT: 12.43 min Scan# 1108
Delta R.T. -0.01 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15




Instrument :
MSVOA_D
ClientSampled :
PN-STONE


Tgt Ion: 95 Resp: 13726
Ion Ratio Lower Upper
95 100
174 83.2 0.0 160.8
176 79.2 0.0 154.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.29 min Scan# 992
Delta R.T. -0.01 min
Lab File: VD060327.D
Acq: 9 Nov 2018 22:15

Tgt Ion: 117 Resp: 26903
Ion Ratio Lower Upper
117 100
82 60.3 43.7 65.5
119 29.9 26.1 39.1





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.38 min Scan# 1204

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :

PN-STONE


Tgt Ion:152 Resp: 17172

Ion Ratio Lower Upper

152 100

115 60.7 24.1 72.3




150 166.2 0.0 310.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 115.00 (114.70 to 115.70): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 22.841 ug/l

RT: 12.43 min Scan# 1108

Delta R.T. -0.18 min

Lab File: VD060327.D


Acq: 9 Nov 2018 22:15

Tgt Ion: 75 Resp: 4692

Ion Ratio Lower Upper



75 100


77 0.0 16.4 49.1#



Abundance Ion 75.00 (74.70 to 75.70): VDC

Ion 76.95 (76.65 to 77.65): VDC





#81

trans-1,4-Dichloro-2-butene

Concen: 100.065 ug/l

RT: 12.43 min Scan# 1108

Delta R.T. 0.08 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15

Instrument :

MSVOA_D

ClientSampleId :

PN-STONE


Tgt Ion: 75 Resp: 5852

Ion Ratio Lower Upper




75 100

53 0.0 68.6 103.0#

89 0.0 34.1 51.1#



Abundance Ion 75.00 (74.70 to 75.70): VDC
Ion 53.00 (52.70 to 53.70): VDC
Ion 89.00 (88.70 to 89.70): VDC



#95

Naphthalene

Concen: 2.114 ug/l

RT: 14.97 min Scan# 1366

Delta R.T. -0.01 min

Lab File: VD060327.D

Acq: 9 Nov 2018 22:15


Tgt Ion: 128 Resp: 994

Ion Ratio Lower Upper

128 100

127 0.0 9.6 14.4#

129 0.0 8.6 12.8#



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

