

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD110918\
 Data File : VD060326.D
 Acq On : 9 Nov 2018 21:48
 Operator : VA/AP
 Sample : J5887-15
 Misc : 4.97g/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 W-3-S

Quant Time: Nov 12 03:39:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.41	168	992051	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	7.45	114	1406304	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.29	117	1018862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.38	152	400117	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.80	65	501060	70.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	141.08%	
35) Dibromofluoromethane	6.33	113	602903	57.49	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	114.98%	
50) Toluene-d8	9.46	98	1564028	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.44	95	477729	42.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.92%	
Target Compounds						
16) Acetone	3.16	43	37058	30.076	ug/l	97
20) Methylene Chloride	3.72	84	102317	9.361	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD110918\
Data File : VD060326.D
Acq On : 9 Nov 2018 21:48
Operator : VA/AP
Sample : J5887-15
Misc : 4.97a/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
W-3-S

Quant Time: Nov 12 03:39:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

