

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110921\
 Data File : VD070954.D
 Acq On : 09 Nov 2021 17:16
 Operator : VA/SY
 Sample : M4556-01RE
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-110521RE

Quant Time: Nov 10 03:57:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	37106	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	67398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	65968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	31274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	25618	64.844	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 129.680%	
35) Dibromofluoromethane	7.914	113	18517	48.689	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.380%	
50) Toluene-d8	10.332	98	83774	56.247	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.500%	
62) 4-Bromofluorobenzene	12.620	95	33318	62.129	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 124.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110921\
Data File : VD070954.D
Acq On : 09 Nov 2021 17:16
Operator : VA/SY
Sample : M4556-01RE
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-110521RE

Quant Time: Nov 10 03:57:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

