

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110921\
 Data File : VD070960.D
 Acq On : 09 Nov 2021 20:06
 Operator : VA/SY
 Sample : M4540-09RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5RE

Quant Time: Nov 10 03:59:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	33024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	64441	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	66080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	30541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	26756	76.096	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.200%
35) Dibromofluoromethane	7.914	113	18240	50.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.320%
50) Toluene-d8	10.332	98	80702	56.671	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.340%
62) 4-Bromofluorobenzene	12.620	95	33471	65.278	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	130.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110921\
Data File : VD070960.D
Acq On : 09 Nov 2021 20:06
Operator : VA/SY
Sample : M4540-09RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5RE

Quant Time: Nov 10 03:59:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

