

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111021\
 Data File : VD070973.D
 Acq On : 10 Nov 2021 16:38
 Operator : VA/SY
 Sample : M4541-01RE
 Misc : 4.37G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-3-ASRE

Quant Time: Nov 11 06:30:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	9314	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.856	114	17593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	16951	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	7307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	7106	71.657	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 143.320%	
35) Dibromofluoromethane	7.897	113	4333	43.647	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 87.300%	
50) Toluene-d8	10.332	98	21213	54.563	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.120%	
62) 4-Bromofluorobenzene	12.620	95	7756	55.406	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.820%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111021\
Data File : VD070973.D
Acq On : 10 Nov 2021 16:38
Operator : VA/SY
Sample : M4541-01RE
Misc : 4.37G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-3-ASRE

Quant Time: Nov 11 06:30:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

