

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
 Data File : VD074872.D
 Acq On : 10 Nov 2022 17:15
 Operator : VA/SY
 Sample : N5520-03RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124037RE

Quant Time: Nov 11 07:18:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	81440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	135850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	135041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	60595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	43956	68.147	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	136.300%	
35) Dibromofluoromethane	7.810	113	50084	62.324	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	124.640%	
50) Toluene-d8	10.269	98	122210	41.691	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	83.380%	
62) 4-Bromofluorobenzene	12.575	95	43761	51.704	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	103.400%	
Target Compounds						
16) Acetone	4.034	43	3502	26.017	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
Data File : VD074872.D
Acq On : 10 Nov 2022 17:15
Operator : VA/SY
Sample : N5520-03RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124037RE

Quant Time: Nov 11 07:18:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

