

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
 Data File : VD074875.D
 Acq On : 10 Nov 2022 18:38
 Operator : VA/SY
 Sample : N5518-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAT-02-110822

Quant Time: Nov 11 07:19:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	41003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	81354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	85007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	37236	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	31225	96.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	192.300%#
35) Dibromofluoromethane	7.804	113	29802	61.927	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.860%
50) Toluene-d8	10.269	98	72005	41.018	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.040%
62) 4-Bromofluorobenzene	12.575	95	26570	52.422	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
Data File : VD074875.D
Acq On : 10 Nov 2022 18:38
Operator : VA/SY
Sample : N5518-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAT-02-110822

Quant Time: Nov 11 07:19:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

