

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
 Data File : VD074870.D
 Acq On : 10 Nov 2022 16:20
 Operator : VA/SY
 Sample : N5436-13RE
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 250FERRYBLVD-TFS-13RE

Quant Time: Nov 10 16:59:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	171284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	261403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	217061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	82727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	54438	40.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.260%
35) Dibromofluoromethane	7.805	113	80472	52.042	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.080%
50) Toluene-d8	10.269	98	208321	36.933	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.860%
62) 4-Bromofluorobenzene	12.575	95	61295	37.637	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.280%


Target Compounds Qvalue


 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111022\
Data File : VD074870.D
Acq On : 10 Nov 2022 16:20
Operator : VA/SY
Sample : N5436-13RE
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
250FERRYBLVD-TFS-13RE

Quant Time: Nov 10 16:59:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20




Instrument :
MSVOA_D
ClientSampleId :
250FERRYBLVD-TFS-13RE

Tgt Ion:168 Resp: 171284
Ion Ratio Lower Upper
168 100
99 49.8 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 40.128 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20

Tgt Ion: 65 Resp: 54438
Ion Ratio Lower Upper
65 100
67 54.2 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074870.D

Acq: 10 Nov 2022 16:20

Instrument :

MSVOA_D

ClientSampleId :

250FERRYBLVD-TFS-13RE





Tgt Ion:114 Resp: 261403

Ion Ratio Lower Upper

114 100

63 17.4 0.0 33.2

88 13.7 0.0 28.2



#35

Dibromofluoromethane

Concen: 52.042 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD074870.D

Acq: 10 Nov 2022 16:20




Tgt Ion:113 Resp: 80472


Ion Ratio Lower Upper

113 100

111 101.8 82.3 123.5

192 19.8 16.4 24.6









#50
Toluene-d8
Concen: 36.933 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20




Instrument :
MSVOA_D
ClientSampleId :
250FERRYBLVD-TFS-13RE


Tgt Ion: 98 Resp: 208321
Ion Ratio Lower Upper
98 100
100 64.3 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 37.637 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20

Tgt Ion: 95 Resp: 61295
Ion Ratio Lower Upper
95 100
174 94.5 0.0 197.0
176 90.6 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20

Instrument :
MSVOA_D
ClientSampleId :
250FERRYBLVD-TFS-13RE

Tgt Ion:	117	Resp:	217061
Ion	Ratio	Lower	Upper
117	100		
82	48.2	41.6	62.4
119	32.2	25.2	37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD074870.D
Acq: 10 Nov 2022 16:20

Tgt Ion:	152	Resp:	82727
Ion	Ratio	Lower	Upper
152	100		
115	54.8	26.8	80.4
150	157.5	0.0	348.2

