

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
 Data File : VD070987.D
 Acq On : 11 Nov 2021 13:06
 Operator : VA/SY
 Sample : M4574-06
 Misc : 5.27G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VWE-B6-110521-29-30

Quant Time: Nov 12 07:59:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	6768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	12574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	11128	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	4697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	4794	66.529	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 133.060%	
35) Dibromofluoromethane	7.914	113	3049	42.972	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 85.940%	
50) Toluene-d8	10.332	98	14149	50.920	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.840%	
62) 4-Bromofluorobenzene	12.620	95	5676	56.732	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 113.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
Data File : VD070987.D
Acq On : 11 Nov 2021 13:06
Operator : VA/SY
Sample : M4574-06
Misc : 5.27G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VWE-B6-110521-29-30

Quant Time: Nov 12 07:59:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

