

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
 Data File : VD071000.D
 Acq On : 11 Nov 2021 19:18
 Operator : VA/SY
 Sample : M4628-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LAMINGTON-COMP

Quant Time: Nov 12 08:04:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	367557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	624360	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	598721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	281536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	175449	44.833	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.660%
35) Dibromofluoromethane	7.908	113	177655	50.425	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.860%
50) Toluene-d8	10.332	98	791259	57.348	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.700%
62) 4-Bromofluorobenzene	12.620	95	289245	58.223	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.440%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	5794	24.844	ug/l #	79
20) Methylene Chloride	4.973	84	19701	4.038	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
Data File : VD071000.D
Acq On : 11 Nov 2021 19:18
Operator : VA/SY
Sample : M4628-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LAMINGTON-COMP

Quant Time: Nov 12 08:04:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

