

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
 Data File : VD071010.D
 Acq On : 12 Nov 2021 03:44
 Operator : VA/SY
 Sample : M4574-06
 Misc : 4.93G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VWE-B6-110521-29-30

Quant Time: Nov 12 08:16:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	325891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	560969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	554713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	254303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	185784	53.543	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.080%	
35) Dibromofluoromethane	7.914	113	169926	53.682	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.360%	
50) Toluene-d8	10.332	98	705775	56.933	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 113.860%	
62) 4-Bromofluorobenzene	12.620	95	269534	60.386	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	25127	44.488	ug/l	90
17) Carbon Disulfide	4.409	76	11524	1.153	ug/l	97
20) Methylene Chloride	4.967	84	26015	6.223	ug/l	94
25) 2-Butanone	7.220	43	9214	27.356	ug/l	94
95) Naphthalene	15.379	128	20915	2.582	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
Data File : VD071010.D
Acq On : 12 Nov 2021 03:44
Operator : VA/SY
Sample : M4574-06
Misc : 4.93G/5.00ml/MSVOA_D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VWE-B6-110521-29-30

Quant Time: Nov 12 08:16:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

