

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
 Data File : VD070981.D
 Acq On : 11 Nov 2021 10:07
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 07:56:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.427	0.444	-4.0	85	0.00
3 P	Chloromethane	0.508	0.557	-9.6	89	0.00
4 C	Vinyl Chloride	0.489	0.611	-24.9#	96	0.00
5 T	Bromomethane	0.333	0.404	-21.3	101	0.00
6 T	Chloroethane	0.314	0.401	-27.7#	97	0.00
7 T	Trichlorofluoromethane	0.787	0.893	-13.5	87	0.00
8 T	Diethyl Ether	0.260	0.269	-3.5	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.485	0.543	-12.0	86	0.00
10 T	Methyl Iodide	0.476	0.553	-16.2	77	0.00
11 T	Tert butyl alcohol	0.039	0.037	5.1	79	0.00
12 CM	1,1-Dichloroethene	0.475	0.502	-5.7#	81	0.00
13 T	Acrolein	0.013	0.010	23.1	71	0.00
14 T	Allyl chloride	0.989	1.034	-4.6	79	0.00
15 T	Acrylonitrile	0.125	0.132	-5.6	79	0.00
16 T	Acetone	0.133	0.165	-24.1	91	0.00
17 T	Carbon Disulfide	1.533	1.551	-1.2	77	0.00
18 T	Methyl Acetate	0.457	0.496	-8.5	83	0.00
19 T	Methyl tert-butyl Ether	1.226	1.302	-6.2	81	0.00
20 T	Methylene Chloride	0.665	0.604	9.2	80	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.563	-3.5	78	0.00
22 T	Diisopropyl ether	1.966	2.054	-4.5	78	0.00
23 T	Vinyl Acetate	0.944	1.029	-9.0	81	0.00
24 P	1,1-Dichloroethane	1.050	1.112	-5.9	80	0.00
25 T	2-Butanone	0.169	0.194	-14.8	84	0.00
26 T	2,2-Dichloropropane	0.886	0.995	-12.3	86	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.655	-5.3	80	0.00
28 T	Bromochloromethane	0.445	0.469	-5.4	78	0.00
29 T	Tetrahydrofuran	0.108	0.114	-5.6	81	0.00
30 C	Chloroform	0.998	1.074	-7.6#	82	0.00
31 T	Cyclohexane	1.095	1.068	2.5	80	0.00
32 T	1,1,1-Trichloroethane	0.857	0.952	-11.1	84	0.00
33 S	1,2-Dichloroethane-d4	0.532	0.524	1.5	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.282	0.284	-0.7	79	0.00
36 T	1,1-Dichloropropene	0.446	0.503	-12.8	83	0.00
37 T	Ethyl Acetate	0.204	0.230	-12.7	82	0.00
38 T	Carbon Tetrachloride	0.411	0.476	-15.8	86	0.00
39 T	Methylcyclohexane	0.577	0.624	-8.1	81	0.00
40 TM	Benzene	1.257	1.368	-8.8	80	0.00
41 T	Methacrylonitrile	0.138	0.131	5.1	78	0.00
42 TM	1,2-Dichloroethane	0.356	0.401	-12.6	82	0.00
43 T	Isopropyl Acetate	0.431	0.469	-8.8	79	0.00
44 TM	Trichloroethene	0.324	0.354	-9.3	81	0.00
45 C	1,2-Dichloropropane	0.340	0.361	-6.2#	79	0.00
46 T	Dibromomethane	0.162	0.183	-13.0	82	0.00
47 T	Bromodichloromethane	0.428	0.479	-11.9	82	0.00
48 T	Methyl methacrylate	0.211	0.234	-10.9	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
 Data File : VD070981.D
 Acq On : 11 Nov 2021 10:07
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 12 07:56:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	83	0.00
50 S	Toluene-d8	1.105	1.082	2.1	77	0.00
51 T	4-Methyl-2-Pentanone	0.209	0.239	-14.4	83	0.00
52 CM	Toluene	0.793	0.857	-8.1#	80	0.00
53 T	t-1,3-Dichloropropene	0.432	0.466	-7.9	80	0.00
54 T	cis-1,3-Dichloropropene	0.509	0.537	-5.5	79	0.00
55 T	1,1,2-Trichloroethane	0.222	0.250	-12.6	83	0.00
56 T	Ethyl methacrylate	0.315	0.341	-8.3	78	0.00
57 T	1,3-Dichloropropane	0.404	0.449	-11.1	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.173	-5.5	77	0.00
59 T	2-Hexanone	0.152	0.176	-15.8	84	0.00
60 T	Dibromochloromethane	0.267	0.301	-12.7	81	0.00
61 T	1,2-Dibromoethane	0.216	0.237	-9.7	81	0.00
62 S	4-Bromofluorobenzene	0.398	0.380	4.5	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.292	0.316	-8.2	81	0.00
65 PM	Chlorobenzene	0.892	0.982	-10.1	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.317	0.353	-11.4	81	0.00
67 C	Ethyl Benzene	1.689	1.866	-10.5#	80	0.00
68 T	m/p-Xylenes	0.631	0.702	-11.3	82	0.00
69 T	o-Xylene	0.601	0.658	-9.5	79	0.00
70 T	Styrene	1.017	1.119	-10.0	80	0.00
71 P	Bromoform	0.165	0.188	-13.9	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.602	3.993	-10.9	82	0.00
74 T	N-amyl acetate	1.029	1.114	-8.3	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.620	0.687	-10.8	81	0.00
76 T	1,2,3-Trichloropropane	0.427	0.438	-2.6	78	-0.01
77 T	Bromobenzene	0.770	0.846	-9.9	82	0.00
78 T	n-propylbenzene	4.463	4.938	-10.6	82	0.00
79 T	2-Chlorotoluene	2.545	2.752	-8.1	80	0.00
80 T	1,3,5-Trimethylbenzene	2.966	3.266	-10.1	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.204	0.224	-9.8	79	0.00
82 T	4-Chlorotoluene	2.626	2.848	-8.5	81	0.00
83 T	tert-Butylbenzene	2.555	2.813	-10.1	82	0.00
84 T	1,2,4-Trimethylbenzene	2.905	3.180	-9.5	81	0.00
85 T	sec-Butylbenzene	3.841	4.274	-11.3	82	0.00
86 T	p-Isopropyltoluene	3.084	3.501	-13.5	84	0.00
87 T	1,3-Dichlorobenzene	1.505	1.668	-10.8	82	0.00
88 T	1,4-Dichlorobenzene	1.507	1.641	-8.9	83	0.00
89 T	n-Butylbenzene	3.099	3.450	-11.3	81	0.00
90 T	Hexachloroethane	0.610	0.669	-9.7	80	0.00
91 T	1,2-Dichlorobenzene	1.304	1.430	-9.7	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.106	-3.9	78	0.00
93 T	1,2,4-Trichlorobenzene	0.834	0.927	-11.2	82	0.00
94 T	Hexachlorobutadiene	0.421	0.497	-18.1	84	0.00
95 T	Naphthalene	1.593	1.778	-11.6	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111121\
Data File : VD070981.D
Acq On : 11 Nov 2021 10:07
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 12 07:56:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.701	0.789	-12.6	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6