

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD111218\
 Data File : VD060330.D
 Acq On : 12 Nov 2018 11:04
 Operator : VA/AP
 Sample : VD1112SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1112SBL01

Quant Time: Nov 13 00:49:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.40	168	1789956	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.43	114	2524255	50.00	ug/l	-0.03
63) Chlorobenzene-d5	11.27	117	2025013	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.37	152	979423	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	720311	56.20	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	112.40%	
35) Dibromofluoromethane	6.32	113	991854	52.69	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	105.38%	
50) Toluene-d8	9.44	98	2903509	52.70	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.41	95	999094	50.05	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	100.10%	
Target Compounds						
20) Methylene Chloride	3.73	84	114321	5.797	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD111218\
Data File : VD060330.D
Acq On : 12 Nov 2018 11:04
Operator : VA/AP
Sample : VD1112SBL01
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1112SBL01

Quant Time: Nov 13 00:49:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

