

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111323\
 Data File : VD077541.D
 Acq On : 13 Nov 2023 11:42
 Operator : JC/SY
 Sample : VIBLK112
 Misc : 5.00G/10ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK112

Quant Time: Nov 16 00:41:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 10 02:07:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	367352	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	330477	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	144458	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.281	65	164542	17.608	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.440%
7) Chloroethane-d5	2.811	69	133596	18.198	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.800%
11) 1,1-Dichloroethene-d2	3.934	65	53080	18.369	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.480%
21) 2-Butanone-d5	6.993	46	57862	35.763	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.520%
24) Chloroform-d	7.569	84	199119	18.764	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	75.040%
26) 1,2-Dichloroethane-d4	8.234	65	110150	19.514	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.040%
32) Benzene-d6	8.205	84	428317	19.969	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.880%
36) 1,2-Dichloropropane-d6	9.210	67	139004	19.896	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.600%
41) Toluene-d8	10.269	98	382835	19.752	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.000%
43) trans-1,3-Dichloroprop...	10.528	79	57279	19.014	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.040%
47) 2-Hexanone-d5	10.881	63	43135	34.513	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.020%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	82358	17.792	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	71.160%
66) 1,2-Dichlorobenzene-d4	13.816	152	102163	19.880	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.520%
Target Compounds						Qvalue
16) Methylene chloride	4.817	84	9257	1.493	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111323\
Data File : VD077541.D
Acq On : 13 Nov 2023 11:42
Operator : JC/SY
Sample : VIBLK112
Misc : 5.00G/10ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK112

Quant Time: Nov 16 00:41:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 10 02:07:44 2023
Response via : Initial Calibration

