

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
 Data File : VD060374.D
 Acq On : 14 Nov 2018 12:48
 Operator : JC/SY
 Sample : J5770-01
 Misc : 6.52g/5ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-01-111218-A

Quant Time: Nov 14 16:17:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.38	168	1491493	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	7.43	114	2054997	50.00	ug/l	-0.03
63) Chlorobenzene-d5	11.27	117	1310412	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.36	152	529767	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.77	65	514253	48.15	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	96.30%	
35) Dibromofluoromethane	6.31	113	696890	45.48	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	90.96%	
50) Toluene-d8	9.44	98	1231568	27.46	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	54.92%	
62) 4-Bromofluorobenzene	12.42	95	508171	31.27	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	62.54%	
Target Compounds						
16) Acetone	3.15	43	84489	45.61	ug/l	98
20) Methylene Chloride	3.71	84	104699	6.37	ug/l	91
25) 2-Butanone	5.60	43	18754	4.87	ug/l #	91
40) Benzene	6.77	78	81603	1.57	ug/l	99
52) Toluene	9.52	92	52777	1.59	ug/l	91
67) Ethyl Benzene	11.41	91	50616	1.04	ug/l	92
68) m/p-Xylenes	11.55	106	32684	1.82	ug/l	95
95) Naphthalene	14.96	128	165118	11.38	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
Data File : VD060374.D
Acq On : 14 Nov 2018 12:48
Operator : JC/SY
Sample : J5770-01
Misc : 6.52g/5ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-01-111218-A

Quant Time: Nov 14 16:17:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

