

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
 Data File : VD060375.D
 Acq On : 14 Nov 2018 13:17
 Operator : JC/SY
 Sample : J5770-03
 Misc : 6.52g/5ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-01-111218-B

Quant Time: Nov 14 16:20:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	1640707	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	7.42	114	2212680	50.00	ug/l	-0.04
63) Chlorobenzene-d5	11.27	117	434411	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.36	152	711488	50.00	ug/l	-0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.76	65	502160	42.75	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	85.50%	
35) Dibromofluoromethane	6.30	113	704938	42.72	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	85.44%	
50) Toluene-d8	9.43	98	245374	5.08	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	10.16%	
62) 4-Bromofluorobenzene	12.41	95	610221	34.87	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	69.74%	
Target Compounds						
16) Acetone	3.15	43	69214	33.97	ug/l	98
20) Methylene Chloride	3.71	84	146455	8.10	ug/l	97
67) Ethyl Benzene	11.41	91	20298	1.26	ug/l	98
68) m/p-Xylenes	11.55	106	11153	1.88	ug/l	82
69) o-Xylene	11.92	106	9681	1.69	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
Data File : VD060375.D
Acq On : 14 Nov 2018 13:17
Operator : JC/SY
Sample : J5770-03
Misc : 6.52g/5ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-01-111218-B

Quant Time: Nov 14 16:20:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

