

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
 Data File : VD060377.D
 Acq On : 14 Nov 2018 14:13
 Operator : JC/SY
 Sample : J5925-09
 Misc : 10.55g/5ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5S1

Quant Time: Nov 14 16:26:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	1426092	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	7.42	114	1937411	50.00	ug/l	-0.04
63) Chlorobenzene-d5	11.27	117	1385435	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.36	152	489070	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.76	65	449182	43.99	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	87.98%	
35) Dibromofluoromethane	6.30	113	529477	36.65	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	73.30%	
50) Toluene-d8	9.44	98	1565857	37.03	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	74.06%	
62) 4-Bromofluorobenzene	12.41	95	434148	28.34	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	56.68%	
Target Compounds						
16) Acetone	3.15	43	88868	50.17	ug/l	# 87
17) Carbon Disulfide	3.30	76	41319	1.84	ug/l	99
20) Methylene Chloride	3.70	84	141378	9.00	ug/l	96
25) 2-Butanone	5.60	43	17725	4.81	ug/l	99
52) Toluene	9.52	92	96643	3.10	ug/l	99
95) Naphthalene	14.96	128	13450	1.00	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
Data File : VD060377.D
Acq On : 14 Nov 2018 14:13
Operator : JC/SY
Sample : J5925-09
Misc : 10.55g/5ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5S1

Quant Time: Nov 14 16:26:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

