

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
 Data File : VD060385.D
 Acq On : 14 Nov 2018 18:06
 Operator : JC/SY
 Sample : J5918-06
 Misc : 5.00g/5ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FDSB-12(26-28)

Quant Time: Nov 15 05:58:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 06 13:56:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.37	168	1441593	50.00	ug/l	-0.05
34) 1,4-Difluorobenzene	7.41	114	2038289	50.00	ug/l	-0.05
63) Chlorobenzene-d5	11.26	117	1687176	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	13.35	152	885906	50.00	ug/l	-0.03

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.76	65	429029	41.56	ug/l	-0.04
Spiked Amount			Recovery	=	83.12%	
35) Dibromofluoromethane	6.30	113	584403	38.45	ug/l	-0.04
Spiked Amount			Recovery	=	76.90%	
50) Toluene-d8	9.42	98	1672470	37.60	ug/l	-0.04
Spiked Amount			Recovery	=	75.20%	
62) 4-Bromofluorobenzene	12.41	95	632122	39.22	ug/l	-0.03
Spiked Amount			Recovery	=	78.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.14	43	33053	18.46	ug/l	# 89
20) Methylene Chloride	3.69	84	83196	5.24	ug/l	94
30) Chloroform	6.07	83	91847	3.56	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD111418\
Data File : VD060385.D
Acq On : 14 Nov 2018 18:06
Operator : JC/SY
Sample : J5918-06
Misc : 5.00g/5ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FDSB-12(26-28)

Quant Time: Nov 15 05:58:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 06 13:56:04 2018
Response via : Initial Calibration

