

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD111418\  
 Data File : VD060370.D  
 Acq On : 14 Nov 2018 10:20  
 Operator : JC/SY  
 Sample : VSTDCCC050  
 Misc : 5.00g/5ml/MSVOA\_D/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 LabSampleId :  
 VSTDCCC050

Quant Time: Nov 14 14:33:38 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D110618S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 06 13:56:04 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0   | 106   | -0.03    |
| 2 T   | Dichlorodifluoromethane     | 0.448 | 0.403 | 10.0  | 105   | 0.00     |
| 3 P   | Chloromethane               | 0.474 | 0.399 | 15.8  | 98    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.387 | 0.356 | 8.0#  | 102   | 0.00     |
| 5 T   | Bromomethane                | 0.092 | 0.095 | -3.3  | 96    | -0.01    |
| 6 T   | Chloroethane                | 0.128 | 0.118 | 7.8   | 92    | -0.01    |
| 7 T   | Trichlorofluoromethane      | 0.456 | 0.465 | -2.0  | 108   | -0.01    |
| 8 T   | Diethyl Ether               | 0.083 | 0.081 | 2.4   | 102   | -0.01    |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.300 | 0.294 | 2.0   | 102   | -0.01    |
| 10 T  | Methyl Iodide               | 0.390 | 0.356 | 8.7   | 96    | -0.01    |
| 11 T  | Tert butyl alcohol          | 0.015 | 0.016 | -6.7  | 102   | -0.03    |
| 12 CM | 1,1-Dichloroethene          | 0.244 | 0.235 | 3.7#  | 100   | -0.02    |
| 13 T  | Acrolein                    | 0.017 | 0.015 | 11.8  | 93    | -0.01    |
| 14 T  | Allyl chloride              | 0.467 | 0.451 | 3.4   | 100   | -0.02    |
| 15 T  | Acrylonitrile               | 0.086 | 0.088 | -2.3  | 105   | -0.02    |
| 16 T  | Acetone                     | 0.062 | 0.065 | -4.8  | 114   | -0.02    |
| 17 T  | Carbon Disulfide            | 0.787 | 0.700 | 11.1  | 95    | -0.01    |
| 18 T  | Methyl Acetate              | 0.132 | 0.122 | 7.6   | 105   | -0.01    |
| 19 T  | Methyl tert-butyl Ether     | 0.731 | 0.732 | -0.1  | 102   | -0.03    |
| 20 T  | Methylene Chloride          | 0.551 | 0.573 | -4.0  | 117   | -0.02    |
| 21 T  | trans-1,2-Dichloroethene    | 0.542 | 0.546 | -0.7  | 105   | -0.02    |
| 22 T  | Diisopropyl ether           | 1.744 | 1.779 | -2.0  | 105   | -0.03    |
| 23 T  | Vinyl Acetate               | 0.868 | 0.877 | -1.0  | 103   | -0.03    |
| 24 P  | 1,1-Dichloroethane          | 0.894 | 0.925 | -3.5  | 107   | -0.03    |
| 25 T  | 2-Butanone                  | 0.129 | 0.136 | -5.4  | 108   | -0.03    |
| 26 T  | 2,2-Dichloropropane         | 0.695 | 0.755 | -8.6  | 113   | -0.03    |
| 27 T  | cis-1,2-Dichloroethene      | 0.564 | 0.584 | -3.5  | 108   | -0.03    |
| 28 T  | Bromochloromethane          | 0.395 | 0.422 | -6.8  | 114   | -0.03    |
| 29    | Tetrahydrofuran             | 0.068 | 0.068 | 0.0   | 104   | -0.03    |
| 30 C  | Chloroform                  | 0.894 | 0.943 | -5.5# | 108   | -0.03    |
| 31 T  | Cyclohexane                 | 0.855 | 0.809 | 5.4   | 103   | -0.04    |
| 32 T  | 1,1,1-Trichloroethane       | 0.753 | 0.766 | -1.7  | 104   | -0.03    |
| 33 S  | 1,2-Dichloroethane-d4       | 0.358 | 0.377 | -5.3  | 108   | -0.03    |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 107   | -0.03    |
| 35 S  | Dibromofluoromethane        | 0.373 | 0.390 | -4.6  | 105   | -0.03    |
| 36 T  | 1,1-Dichloropropene         | 0.494 | 0.505 | -2.2  | 109   | -0.04    |
| 37 T  | Ethyl Acetate               | 0.227 | 0.237 | -4.4  | 107   | -0.02    |
| 38 T  | Carbon Tetrachloride        | 0.458 | 0.463 | -1.1  | 107   | -0.03    |
| 39 T  | Methylcyclohexane           | 0.568 | 0.581 | -2.3  | 108   | -0.03    |
| 40 TM | Benzene                     | 1.261 | 1.345 | -6.7  | 111   | -0.03    |
| 41 T  | Methacrylonitrile           | 0.260 | 0.261 | -0.4  | 105   | -0.03    |
| 42 TM | 1,2-Dichloroethane          | 0.330 | 0.349 | -5.8  | 109   | -0.03    |
| 43 T  | Isopropyl Acetate           | 0.286 | 0.292 | -2.1  | 104   | -0.03    |
| 44 TM | Trichloroethene             | 0.382 | 0.407 | -6.5  | 110   | -0.03    |
| 45 C  | 1,2-Dichloropropane         | 0.318 | 0.334 | -5.0# | 108   | -0.03    |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD111418\  
 Data File : VD060370.D  
 Acq On : 14 Nov 2018 10:20  
 Operator : JC/SY  
 Sample : VSTDCCC050  
 Misc : 5.00g/5ml/MSVOA\_D/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 LabSampleId :  
 VSTDCCC050

Quant Time: Nov 14 14:33:38 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D110618S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 06 13:56:04 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 46 T  | Dibromomethane              | 0.186 | 0.197 | -5.9  | 109   | -0.03    |
| 47 T  | Bromodichloromethane        | 0.439 | 0.465 | -5.9  | 107   | -0.03    |
| 48 T  | Methyl methacrylate         | 0.173 | 0.178 | -2.9  | 105   | -0.03    |
| 49 T  | 1,4-Dioxane                 | 0.002 | 0.002 | 0.0   | 111   | -0.03    |
| 50 S  | Toluene-d8                  | 1.091 | 1.121 | -2.7  | 104   | -0.03    |
| 51 T  | 4-Methyl-2-Pentanone        | 0.186 | 0.193 | -3.8  | 109   | -0.03    |
| 52 CM | Toluene                     | 0.805 | 0.823 | -2.2# | 107   | -0.03    |
| 53 T  | t-1,3-Dichloropropene       | 0.371 | 0.401 | -8.1  | 108   | -0.03    |
| 54 T  | cis-1,3-Dichloropropene     | 0.476 | 0.511 | -7.4  | 110   | -0.03    |
| 55 T  | 1,1,2-Trichloroethane       | 0.233 | 0.235 | -0.9  | 110   | -0.03    |
| 56 T  | Ethyl methacrylate          | 0.215 | 0.233 | -8.4  | 106   | -0.03    |
| 57 T  | 1,3-Dichloropropane         | 0.356 | 0.382 | -7.3  | 111   | -0.03    |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.130 | 0.133 | -2.3  | 106   | -0.03    |
| 59 T  | 2-Hexanone                  | 0.132 | 0.147 | -11.4 | 112   | -0.03    |
| 60 T  | Dibromochloromethane        | 0.284 | 0.312 | -9.9  | 108   | -0.02    |
| 61 T  | 1,2-Dibromoethane           | 0.234 | 0.246 | -5.1  | 106   | -0.02    |
| 62 S  | 4-Bromofluorobenzene        | 0.395 | 0.409 | -3.5  | 108   | -0.02    |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 108   | -0.03    |
| 64 T  | Tetrachloroethene           | 0.446 | 0.473 | -6.1  | 112   | -0.03    |
| 65 PM | Chlorobenzene               | 1.083 | 1.113 | -2.8  | 109   | -0.03    |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.354 | 0.384 | -8.5  | 115   | -0.02    |
| 67 C  | Ethyl Benzene               | 1.853 | 1.849 | 0.2#  | 110   | -0.02    |
| 68 T  | m/p-Xylenes                 | 0.684 | 0.694 | -1.5  | 107   | -0.02    |
| 69 T  | o-Xylene                    | 0.660 | 0.684 | -3.6  | 108   | -0.02    |
| 70 T  | Styrene                     | 1.067 | 1.095 | -2.6  | 110   | -0.02    |
| 71 P  | Bromoform                   | 0.238 | 0.259 | -8.8  | 109   | -0.02    |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 110   | -0.02    |
| 73 T  | Isopropylbenzene            | 3.643 | 3.549 | 2.6   | 111   | -0.02    |
| 74 T  | N-amyl acetate              | 0.934 | 0.893 | 4.4   | 101   | -0.02    |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 0.596 | 0.588 | 1.3   | 107   | -0.01    |
| 76 T  | 1,2,3-Trichloropropane      | 0.598 | 0.603 | -0.8  | 109   | -0.02    |
| 77 T  | Bromobenzene                | 0.968 | 0.982 | -1.4  | 112   | -0.02    |
| 78 T  | n-propylbenzene             | 4.547 | 4.487 | 1.3   | 108   | -0.02    |
| 79 T  | 2-Chlorotoluene             | 2.490 | 2.482 | 0.3   | 112   | -0.02    |
| 80 T  | 1,3,5-Trimethylbenzene      | 2.831 | 2.698 | 4.7   | 106   | -0.02    |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.170 | 0.178 | -4.7  | 110   | -0.02    |
| 82 T  | 4-Chlorotoluene             | 2.797 | 2.710 | 3.1   | 110   | -0.02    |
| 83 T  | tert-Butylbenzene           | 3.143 | 3.234 | -2.9  | 112   | -0.02    |
| 84 T  | 1,2,4-Trimethylbenzene      | 2.827 | 2.773 | 1.9   | 106   | -0.01    |
| 85 T  | sec-Butylbenzene            | 3.856 | 3.829 | 0.7   | 109   | -0.01    |
| 86 T  | p-Isopropyltoluene          | 3.105 | 2.983 | 3.9   | 104   | -0.02    |
| 87 T  | 1,3-Dichlorobenzene         | 1.722 | 1.750 | -1.6  | 109   | -0.02    |
| 88 T  | 1,4-Dichlorobenzene         | 1.681 | 1.705 | -1.4  | 110   | -0.02    |
| 89 T  | n-Butylbenzene              | 3.168 | 3.120 | 1.5   | 109   | -0.02    |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD111418\  
Data File : VD060370.D  
Acq On : 14 Nov 2018 10:20  
Operator : JC/SY  
Sample : VSTDCCC050  
Misc : 5.00g/5ml/MSVOA\_D/SOIL  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
LabSampleId :  
VSTDCCC050

Quant Time: Nov 14 14:33:38 2018  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D110618S.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 06 13:56:04 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 90 T | Hexachloroethane            | 0.736 | 0.765 | -3.9 | 112   | -0.02    |
| 91 T | 1,2-Dichlorobenzene         | 1.414 | 1.402 | 0.8  | 109   | -0.02    |
| 92 T | 1,2-Dibromo-3-Chloropropane | 0.072 | 0.072 | 0.0  | 104   | -0.01    |
| 93 T | 1,2,4-Trichlorobenzene      | 1.121 | 1.110 | 1.0  | 106   | -0.02    |
| 94 T | Hexachlorobutadiene         | 0.839 | 0.809 | 3.6  | 103   | -0.02    |
| 95 T | Naphthalene                 | 1.369 | 1.353 | 1.2  | 105   | -0.02    |
| 96 T | 1,2,3-Trichlorobenzene      | 0.898 | 0.890 | 0.9  | 107   | -0.02    |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6