


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111419\
 Data File : VD064207.D
 Acq On : 14 Nov 2019 09:36
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Title : SW846 8260
 Last Update : Tue Nov 12 02:43:24 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	32797	PASS
75	95	30	60	55.5	83050	PASS
95	95	100	100	100.0	149632	PASS
96	95	5	9	6.6	9912	PASS
173	174	0.00	2	1.4	1951	PASS
174	95	50	100	96.5	144448	PASS
175	174	5	9	7.6	10979	PASS
176	174	95	101	100.2	144704	PASS
177	176	5	9	7.7	11081	PASS