

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
 Data File : VD074929.D
 Acq On : 14 Nov 2022 12:45
 Operator : VA/SY
 Sample : N5545-01RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-1-110922RE

Quant Time: Nov 15 00:16:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	5104	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	5042	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	2477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	2745	192.111	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 384.220%#	
35) Dibromofluoromethane	7.811	113	2031	64.553	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 129.100%	
50) Toluene-d8	10.275	98	6315	60.932	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 121.860%	
62) 4-Bromofluorobenzene	12.569	95	2215	69.601	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 139.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
Data File : VD074929.D
Acq On : 14 Nov 2022 12:45
Operator : VA/SY
Sample : N5545-01RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-1-110922RE

Quant Time: Nov 15 00:16:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

