

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
 Data File : VD074930.D
 Acq On : 14 Nov 2022 13:12
 Operator : VA/SY
 Sample : N5607-01RE
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01RE

Quant Time: Nov 15 00:23:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	1783	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	3495	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	3384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	1177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	1667	152.899	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 305.800%#	
35) Dibromofluoromethane	7.810	113	1349	62.615	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 125.240%	
50) Toluene-d8	10.269	98	3815	53.757	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.520%	
62) 4-Bromofluorobenzene	12.569	95	1201	55.112	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
Data File : VD074930.D
Acq On : 14 Nov 2022 13:12
Operator : VA/SY
Sample : N5607-01RE
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01RE

Quant Time: Nov 15 00:23:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

