

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
 Data File : VD074933.D
 Acq On : 14 Nov 2022 14:34
 Operator : VA/SY
 Sample : N5607-07RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-38RE

Quant Time: Nov 15 00:16:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	25924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	50567	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	49452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	18795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	24204	152.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 305.360%#	
35) Dibromofluoromethane	7.805	113	21357	68.516	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 137.040%	
50) Toluene-d8	10.269	98	55452	54.005	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.020%	
62) 4-Bromofluorobenzene	12.575	95	16726	53.049	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.100%	
Target Compounds						Qvalue
16) Acetone	4.022	43	1592	37.200	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
Data File : VD074933.D
Acq On : 14 Nov 2022 14:34
Operator : VA/SY
Sample : N5607-07RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-38RE

Quant Time: Nov 15 00:16:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

