

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
 Data File : VD074926.D
 Acq On : 14 Nov 2022 11:23
 Operator : VA/SY
 Sample : VD1114SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1114SBL01

Quant Time: Nov 15 00:15:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	141934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	218037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	194507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	80803	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	49933	53.342	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.680%
35) Dibromofluoromethane	7.810	113	70692	52.596	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.200%
50) Toluene-d8	10.269	98	206063	46.543	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.080%
62) 4-Bromofluorobenzene	12.575	95	61232	45.040	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111422\
Data File : VD074926.D
Acq On : 14 Nov 2022 11:23
Operator : VA/SY
Sample : VD1114SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1114SBL01

Quant Time: Nov 15 00:15:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23




Instrument :
MSVOA_D
ClientSampleId :
VD1114SBL01

Tgt Ion:168 Resp: 141934
Ion Ratio Lower Upper
168 100
99 50.7 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 53.342 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23

Tgt Ion: 65 Resp: 49933
Ion Ratio Lower Upper
65 100
67 55.7 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074926.D

Acq: 14 Nov 2022 11:23

Instrument :

MSVOA_D

ClientSampleId :

VD1114SBL01




Tgt Ion:114 Resp: 218037

Ion Ratio Lower Upper

114 100

63 16.5 0.0 33.2

88 13.5 0.0 28.2



#35

Dibromofluoromethane

Concen: 52.596 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.005 min

Lab File: VD074926.D

Acq: 14 Nov 2022 11:23




Tgt Ion:113 Resp: 70692


Ion Ratio Lower Upper

113 100

111 103.5 82.3 123.5

192 19.6 16.4 24.6








#50
Toluene-d8
Concen: 46.543 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23




Instrument :
MSVOA_D
ClientSampleId :
VD1114SBL01

Tgt Ion: 98 Resp: 206063
Ion Ratio Lower Upper
98 100
100 64.4 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 45.040 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23

Tgt Ion: 95 Resp: 61232
Ion Ratio Lower Upper
95 100
174 94.2 0.0 197.0
176 87.7 0.0 182.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.587 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23

Instrument :
MSVOA_D
ClientSampleId :
VD1114SBL01

Tgt Ion:117 Resp: 194507
Ion Ratio Lower Upper
117 100
82 50.1 41.6 62.4
119 32.4 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. 0.006 min
Lab File: VD074926.D
Acq: 14 Nov 2022 11:23

Tgt Ion:152 Resp: 80803
Ion Ratio Lower Upper
152 100
115 56.5 26.8 80.4
150 159.1 0.0 348.2

