



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111423\
Data File : VD077552.D
Acq On : 14 Nov 2023 08:51
Operator : JC/SY
Sample : BFB177
Misc : 5.00G/10ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB177

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Title : SFAM01.0
Last Update : Fri Nov 10 02:07:44 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.5	24515	PASS
75	95	30	80	57.6	60173	PASS
95	95	100	100	100.0	104520	PASS
96	95	5	9	6.9	7224	PASS
173	174	0.00	2	0.6	464	PASS
174	95	50	120	80.2	83805	PASS
175	174	5	9	8.0	6675	PASS
176	174	95	105	96.5	80896	PASS
177	176	5	10	6.5	5247	PASS

m/z	Abundance
38.20	165.0
39.90	2052.0
43.90	893.0
46.80	155.0
75.60	174.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177

m/z	Abundance
36.10	1286.0
37.10	6670.0
38.00	6471.0
39.00	2392.0
39.90	1862.0
42.90	156.0
43.90	1927.0
45.10	1356.0
45.80	214.0
47.10	1580.0
47.90	778.0
49.00	5561.0
50.00	22112.0
51.00	7586.0
55.00	359.0
55.90	1497.0
57.00	3844.0
57.90	159.0
59.10	354.0
59.90	1291.0
61.00	5457.0
62.00	6012.0
63.00	4634.0
64.00	426.0
65.00	286.0
66.60	218.0
67.10	394.0
68.00	12889.0
69.00	12045.0
69.90	951.0
70.80	163.0
71.90	831.0
73.00	5881.0
74.00	20488.0
75.00	57848.0
76.00	4582.0
77.00	604.0
78.00	763.0
78.90	4700.0
79.90	1122.0
80.90	4398.0
82.00	1395.0
82.80	176.0
86.10	169.0
86.90	3886.0
87.90	4411.0
90.70	328.0
92.00	3727.0
92.90	4633.0
94.00	13607.0
95.00	95320.0
96.00	6849.0
97.00	277.0
103.00	180.0
103.90	496.0
105.70	404.0
106.00	441.0
106.80	206.0
110.80	255.0
112.80	165.0
115.00	276.0
116.00	531.0
116.90	706.0
117.70	416.0
118.90	555.0
127.80	349.0
129.70	290.0
132.90	283.0
134.70	166.0
136.90	158.0
140.90	1380.0
143.00	1567.0
144.80	152.0
145.90	266.0
147.80	248.0
155.00	190.0
156.80	206.0
158.70	238.0
160.80	215.0
170.50	198.0
173.10	1159.0
173.90	74240.0
175.00	6056.0
175.90	71720.0
176.90	4273.0
178.00	238.0
190.90	275.0
192.80	408.0
207.20	286.0
248.90	284.0
281.00	2009.0
282.10	416.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177

282.80

342.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177

m/z	Abundance
36.10	1242.0
37.00	7368.0
38.00	6486.0
39.10	3089.0
40.00	2170.0
43.20	172.0
44.00	2092.0
45.00	1513.0
45.90	372.0
47.00	2160.0
47.90	1164.0
49.00	6261.0
50.00	27992.0
51.00	8150.0
51.90	549.0
55.00	381.0
56.00	2141.0
57.00	3851.0
57.90	202.0
59.10	160.0
60.00	1330.0
61.00	6075.0
62.00	5659.0
63.00	5795.0
64.00	640.0
65.10	204.0
67.10	515.0
68.00	13737.0
69.00	13095.0
70.00	1047.0
72.10	902.0
73.00	6617.0
74.00	22776.0
75.00	64392.0
76.00	5322.0
76.90	887.0
77.90	741.0
78.90	5514.0
80.00	1838.0
80.90	5695.0
81.90	1292.0
82.90	264.0
85.90	289.0
87.00	5683.0
87.90	5372.0
88.80	172.0
90.90	808.0
92.00	3826.0
93.00	5994.0
94.00	15993.0
95.00	114856.0
96.00	8430.0
97.00	268.0
103.90	602.0
105.00	288.0
106.00	556.0
106.80	263.0
110.80	257.0
112.60	508.0
114.80	201.0
116.00	629.0
116.90	1066.0
117.90	848.0
118.80	646.0
127.90	440.0
128.70	268.0
129.90	466.0
130.90	254.0
132.90	673.0
134.00	203.0
134.90	319.0
136.80	400.0
140.90	1526.0
141.90	246.0
142.90	1434.0
144.10	244.0
145.80	285.0
146.80	173.0
148.10	258.0
149.80	161.0
153.10	239.0
155.10	288.0
156.90	158.0
158.90	295.0
160.80	277.0
171.20	199.0
171.90	542.0
172.80	1392.0
173.90	92088.0
174.90	7257.0
175.90	88736.0
176.90	6231.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177

178.10	373.0
192.80	595.0
195.00	179.0
206.90	250.0
248.90	252.0
265.20	218.0
281.00	2745.0
282.00	878.0
283.00	550.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177

m/z	Abundance
36.00	1150.0
37.00	6209.0
38.00	5649.0
39.00	2216.0
39.90	2041.0
41.10	180.0
42.00	156.0
44.00	2243.0
45.00	1407.0
47.00	1985.0
48.10	794.0
49.00	5394.0
50.00	23440.0
51.00	7370.0
52.00	507.0
55.00	441.0
56.00	2063.0
57.00	3521.0
57.80	248.0
60.00	1302.0
61.00	6284.0
62.00	5746.0
63.00	4605.0
64.00	567.0
64.90	250.0
67.10	289.0
68.00	12690.0
69.00	13348.0
70.00	1078.0
71.00	166.0
72.00	881.0
73.00	6372.0
74.00	20120.0
75.00	58280.0
76.00	4747.0
76.90	689.0
78.00	543.0
78.90	5560.0
80.00	1325.0
80.90	5343.0
82.00	1168.0
83.00	265.0
85.70	159.0
86.90	4611.0
88.00	4881.0
90.90	767.0
92.00	3154.0
92.90	4715.0
94.00	12674.0
95.00	103384.0
96.00	6392.0
96.80	372.0
102.90	177.0
103.90	695.0
105.00	293.0
105.90	727.0
111.00	440.0
111.70	170.0
112.80	196.0
115.10	165.0
116.00	562.0
116.90	991.0
117.80	441.0
118.80	664.0
125.30	193.0
127.90	293.0
128.80	174.0
129.70	583.0
133.00	475.0
134.70	201.0
137.00	222.0
140.90	1281.0
141.80	254.0
142.90	1522.0
147.00	285.0
147.80	160.0
156.80	209.0
171.80	414.0
173.90	85088.0
174.90	6711.0
175.90	82232.0
176.90	5237.0
190.80	604.0
192.80	360.0
193.90	180.0
207.00	514.0
249.10	330.0
264.90	373.0
281.00	2615.0
281.90	1076.0
283.00	695.0

Instrument :
MSVOA_D
ClientSampleId :
BFB177