

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD111519\
 Data File : VD064229.D
 Acq On : 15 Nov 2019 16:04
 Operator : VA/SY
 Sample : K5858-01
 Misc : 5.06G/5.00ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DRUM-A

Quant Time: Nov 16 05:56:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 12 02:43:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	174832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	227972	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	76624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	18628	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	92043	64.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.02%	
35) Dibromofluoromethane	7.91	113	99093	75.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	151.44%	
50) Toluene-d8	10.34	98	163679	33.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.32%	
62) 4-Bromofluorobenzene	12.64	95	22093	12.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	25.84%	
Target Compounds						
20) Methylene Chloride	4.96	84	18108	10.524	ug/l	94
30) Chloroform	7.71	83	13048	4.559	ug/l	98
84) 1,2,4-Trimethylbenzene	13.27	105	4793	4.321	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD111519\
Data File : VD064229.D
Acq On : 15 Nov 2019 16:04
Operator : VA/SY
Sample : K5858-01
Misc : 5.06G/5.00ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DRUM-A

Quant Time: Nov 16 05:56:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 12 02:43:24 2019
Response via : Initial Calibration

