

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111521\
 Data File : VD071029.D
 Acq On : 15 Nov 2021 15:44
 Operator : VA/SY
 Sample : M4603-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ISCO-SOIL-IDW-11082021

Quant Time: Nov 16 03:21:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	249126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	445820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	408099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	174251	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	134957	50.880	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.760%	
35) Dibromofluoromethane	7.914	113	138114	54.901	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.800%	
50) Toluene-d8	10.338	98	544438	55.262	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.520%	
62) 4-Bromofluorobenzene	12.626	95	186988	52.713	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.420%	
Target Compounds						Qvalue
16) Acetone	4.156	43	2539	23.060	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111521\
Data File : VD071029.D
Acq On : 15 Nov 2021 15:44
Operator : VA/SY
Sample : M4603-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ISCO-SOIL-IDW-11082021

Quant Time: Nov 16 03:21:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

