

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111521\
 Data File : VD071030.D
 Acq On : 15 Nov 2021 16:12
 Operator : VA/SY
 Sample : M4651-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-02-111121

Quant Time: Nov 16 03:21:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 10 03:55:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	19313	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	36051	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	38340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	18900	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	15787	76.775	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.560%
35) Dibromofluoromethane	7.908	113	10604	52.126	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.260%
50) Toluene-d8	10.337	98	45267	56.820	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.640%
62) 4-Bromofluorobenzene	12.625	95	18432	64.256	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111521\
Data File : VD071030.D
Acq On : 15 Nov 2021 16:12
Operator : VA/SY
Sample : M4651-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-02-111121

Quant Time: Nov 16 03:21:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110921S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 10 03:55:51 2021
Response via : Initial Calibration

