

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111523\
 Data File : VD077575.D
 Acq On : 15 Nov 2023 09:16
 Operator : JC/SY
 Sample : VD1115SBL01
 Misc : 5.00G/10ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VBLK160

Quant Time: Nov 16 05:42:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 10 02:07:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	296932	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	271145	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	117815	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.281	65	162656	21.534	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.120%
7) Chloroethane-d5	2.811	69	140828	23.733	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.920%
11) 1,1-Dichloroethene-d2	3.934	65	49942	21.382	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.520%
21) 2-Butanone-d5	6.993	46	64702	49.475	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.940%
24) Chloroform-d	7.569	84	199903	23.305	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.240%
26) 1,2-Dichloroethane-d4	8.234	65	112518	24.661	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.640%
32) Benzene-d6	8.205	84	417458	23.722	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.880%
36) 1,2-Dichloropropane-d6	9.210	67	137569	23.999	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.000%
41) Toluene-d8	10.269	98	374792	23.568	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.280%
43) trans-1,3-Dichloroprop...	10.528	79	56472	22.849	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.400%
47) 2-Hexanone-d5	10.881	63	50398	49.148	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.300%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	92838	24.444	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.760%
66) 1,2-Dichlorobenzene-d4	13.816	152	109698	26.173	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111523\
Data File : VD077575.D
Acq On : 15 Nov 2023 09:16
Operator : JC/SY
Sample : VD1115SBL01
Misc : 5.00G/10ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VBLK160

Quant Time: Nov 16 05:42:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 10 02:07:44 2023
Response via : Initial Calibration

