

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074950.D
 Acq On : 16 Nov 2022 15:53
 Operator : VA/SY
 Sample : N5642-05
 Misc : 6.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RIGHT-SIDE-FLATS-25

Quant Time: Nov 17 00:21:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	8169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	8015	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	3109	131.867	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 263.740%#	
35) Dibromofluoromethane	7.804	113	3115	61.859	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 123.720%	
50) Toluene-d8	10.269	98	8929	53.829	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.660%	
62) 4-Bromofluorobenzene	12.575	95	2654	52.106	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
Data File : VD074950.D
Acq On : 16 Nov 2022 15:53
Operator : VA/SY
Sample : N5642-05
Misc : 6.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RIGHT-SIDE-FLATS-25

Quant Time: Nov 17 00:21:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

