

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074954.D
 Acq On : 16 Nov 2022 17:42
 Operator : VA/SY
 Sample : N5642-09
 Misc : 6.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LEFT-SIDE-FLATS-29

Quant Time: Nov 17 00:21:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	138303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	253211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	234124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	99384	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66506	75.377	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.760%
35) Dibromofluoromethane	7.810	113	83444	53.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.920%
50) Toluene-d8	10.275	98	310142	60.320	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	120.640%
62) 4-Bromofluorobenzene	12.575	95	85776	54.330	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
Data File : VD074954.D
Acq On : 16 Nov 2022 17:42
Operator : VA/SY
Sample : N5642-09
Misc : 6.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-29

Quant Time: Nov 17 00:21:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

