

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074945.D
 Acq On : 16 Nov 2022 13:29
 Operator : VA/SY
 Sample : VD1116SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBS01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022

Quant Time: Nov 17 00:20:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	142365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	243860	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	220413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	102758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	50088	53.346	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 106.700%			
35) Dibromofluoromethane	7.810	113	63514	42.252	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 84.500%			
50) Toluene-d8	10.269	98	232588	46.971	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 93.940%			
62) 4-Bromofluorobenzene	12.575	95	76880	50.562	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 101.120%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.934	85	24676	24.665	ug/l	97
3) Chloromethane	2.146	50	36833	22.637	ug/l	94
4) Vinyl Chloride	2.281	62	45462	22.774	ug/l	99
5) Bromomethane	2.687	94	38033	24.875	ug/l	99
6) Chloroethane	2.834	64	33149	22.613	ug/l	99
7) Trichlorofluoromethane	3.175	101	53278	23.424	ug/l	95
8) Diethyl Ether	3.593	74	14004	21.475	ug/l	95
9) 1,1,2-Trichlorotrifluo...	3.969	101	33962	23.399	ug/l	98
10) Methyl Iodide	4.169	142	34640	22.923	ug/l	99
11) Tert butyl alcohol	5.052	59	7621	100.918	ug/l	98
12) 1,1-Dichloroethene	3.940	96	32067	22.681	ug/l	99
13) Acrolein	3.799	56	6129	142.114	ug/l	97
14) Allyl chloride	4.563	41	34335	22.704	ug/l	99
15) Acrylonitrile	5.263	53	27989	109.829	ug/l	98
16) Acetone	4.022	43	24151	102.761	ug/l	95
17) Carbon Disulfide	4.275	76	87681	21.801	ug/l	100
18) Methyl Acetate	4.563	43	15298	23.849	ug/l	96
19) Methyl tert-butyl Ether	5.316	73	59145	20.895	ug/l	98
20) Methylene Chloride	4.799	84	56454	20.463	ug/l	94
21) trans-1,2-Dichloroethene	5.316	96	35742	22.293	ug/l	91
22) Diisopropyl ether	6.222	45	75716	22.463	ug/l #	94
23) Vinyl Acetate	6.163	43	149296m	103.199	ug/l	
24) 1,1-Dichloroethane	6.116	63	57748	22.942	ug/l	100
25) 2-Butanone	7.087	43	28712	93.424	ug/l	97
26) 2,2-Dichloropropane	7.081	77	46165	19.516	ug/l	98
27) cis-1,2-Dichloroethene	7.081	96	34929	19.638	ug/l	98
28) Bromochloromethane	7.428	49	16174	19.917	ug/l	90
29) Tetrahydrofuran	7.452	42	16221	95.353	ug/l	94
30) Chloroform	7.593	83	54040	19.603	ug/l	99
31) Cyclohexane	7.881	56	40267	19.401	ug/l	95
32) 1,1,1-Trichloroethane	7.793	97	47573	20.024	ug/l	97
36) 1,1-Dichloropropene	8.010	75	40188	18.191	ug/l	99
37) Ethyl Acetate	7.169	43	11947	16.116	ug/l	97
38) Carbon Tetrachloride	7.999	117	37281	17.927	ug/l	98
39) Methylcyclohexane	9.275	83	46938	19.832	ug/l	97
40) Benzene	8.257	78	120014	18.892	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074945.D
 Acq On : 16 Nov 2022 13:29
 Operator : VA/SY
 Sample : VD1116SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBS01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022

Quant Time: Nov 17 00:20:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.410	41	6877	17.491	ug/l	91
42) 1,2-Dichloroethane	8.328	62	27306	17.961	ug/l	98
43) Isopropyl Acetate	8.363	43	22853	17.503	ug/l	98
44) Trichloroethene	9.028	130	33317	19.851	ug/l	98
45) 1,2-Dichloropropane	9.304	63	28449	19.931	ug/l	92
46) Dibromomethane	9.393	93	15683	19.619	ug/l	91
47) Bromodichloromethane	9.587	83	38857	19.586	ug/l	97
48) Methyl methacrylate	9.381	41	10637	18.391	ug/l	97
49) 1,4-Dioxane	9.381	88	2930	373.534	ug/l	91
51) 4-Methyl-2-Pentanone	10.163	43	58864	94.567	ug/l	97
52) Toluene	10.334	92	77132	20.307	ug/l	94
53) t-1,3-Dichloropropene	10.557	75	34045	19.426	ug/l	100
54) cis-1,3-Dichloropropene	10.016	75	42730	19.881	ug/l	99
55) 1,1,2-Trichloroethane	10.734	97	22364	20.120	ug/l	98
56) Ethyl methacrylate	10.598	69	22349	18.481	ug/l	97
57) 1,3-Dichloropropane	10.881	76	34583	19.353	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.875	63	38964	96.462	ug/l	98
59) 2-Hexanone	10.922	43	41343	95.277	ug/l	98
60) Dibromochloromethane	11.075	129	26255	19.911	ug/l	98
61) 1,2-Dibromoethane	11.181	107	20773	19.983	ug/l	99
64) Tetrachloroethene	10.810	164	27941	18.907	ug/l	89
65) Chlorobenzene	11.610	112	84202	19.528	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.681	131	28945	19.159	ug/l	99
67) Ethyl Benzene	11.687	91	146199	19.462	ug/l	97
68) m/p-Xylenes	11.792	106	118174	39.455	ug/l	98
69) o-Xylene	12.128	106	53376	19.370	ug/l	96
70) Styrene	12.140	104	91261	19.320	ug/l	98
71) Bromoform	12.304	173	14104	18.039	ug/l #	96
73) Isopropylbenzene	12.422	105	143598	20.303	ug/l	100
74) N-amyl acetate	12.234	43	21080	18.115	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.675	83	24711	19.752	ug/l	97
76) 1,2,3-Trichloropropane	12.728	75	17020m	19.803	ug/l	
77) Bromobenzene	12.704	156	32145	19.421	ug/l	95
78) n-propylbenzene	12.769	91	178820	20.568	ug/l	98
79) 2-Chlorotoluene	12.851	91	96218	20.540	ug/l	99
80) 1,3,5-Trimethylbenzene	12.904	105	118503	20.108	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.469	75	6206	17.763	ug/l	96
82) 4-Chlorotoluene	12.951	91	99712	20.440	ug/l	97
83) tert-Butylbenzene	13.169	119	100274	19.609	ug/l	98
84) 1,2,4-Trimethylbenzene	13.216	105	117901	20.228	ug/l	100
85) sec-Butylbenzene	13.345	105	159916	20.559	ug/l	100
86) p-Isopropyltoluene	13.463	119	131595	20.346	ug/l	99
87) 1,3-Dichlorobenzene	13.457	146	66989	19.457	ug/l	99
88) 1,4-Dichlorobenzene	13.539	146	65619	19.203	ug/l	97
89) n-Butylbenzene	13.792	91	122614	20.419	ug/l	98
90) Hexachloroethane	14.057	117	24119	20.821	ug/l	91
91) 1,2-Dichlorobenzene	13.834	146	58195	19.876	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.445	75	2930	16.987	ug/l	93
93) 1,2,4-Trichlorobenzene	15.104	180	32370	18.163	ug/l	95
94) Hexachlorobutadiene	15.210	225	18666	20.305	ug/l	100
95) Naphthalene	15.333	128	58845	17.789	ug/l	98
96) 1,2,3-Trichlorobenzene	15.528	180	29461	19.116	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
Data File : VD074945.D
Acq On : 16 Nov 2022 13:29
Operator : VA/SY
Sample : VD1116SBSD01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022

Quant Time: Nov 17 00:20:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074945.D
 Acq On : 16 Nov 2022 13:29
 Operator : VA/SY
 Sample : VD1116SBSD01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBSD01

Manual Integrations APPROVED

Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022

Quant Time: Nov 17 00:20:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1069
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01

Tgt Ion:168 Resp: 14236
Ion Ratio Lower Upper
168 100
99 53.7 40.9 61.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#2
Dichlorodifluoromethane
Concen: 24.665 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 85 Resp: 24676
Ion Ratio Lower Upper
85 100
87 33.3 17.5 52.6





#3
Chloromethane
Concen: 22.637 ug/l
RT: 2.146 min Scan# 95
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 50 Resp: 3683
Ion Ratio Lower Upper
50 100
52 33.7 30.0 45.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#4
Vinyl Chloride
Concen: 22.774 ug/l
RT: 2.281 min Scan# 118
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 62 Resp: 45462
Ion Ratio Lower Upper
62 100
64 31.4 25.5 38.3





#5
Bromomethane
Concen: 24.875 ug/l
RT: 2.687 min Scan# 187
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01




Tgt Ion: 94 Resp: 3803
Ion Ratio Lower Upper
94 100
96 94.4 74.6 112.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#6
Chloroethane
Concen: 22.613 ug/l
RT: 2.834 min Scan# 212
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29



Tgt Ion: 64 Resp: 33149
Ion Ratio Lower Upper
64 100
66 33.1 25.8 38.8









#7
Trichlorofluoromethane
Concen: 23.424 ug/l
RT: 3.175 min Scan# 21
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01

Tgt Ion:101 Resp: 53278
Ion Ratio Lower Upper
101 100
103 66.5 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#8
Diethyl Ether
Concen: 21.475 ug/l
RT: 3.593 min Scan# 341
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 74 Resp: 14004
Ion Ratio Lower Upper
74 100
45 79.8 37.9 113.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 23.399 ug/l
RT: 3.969 min Scan# 405
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:101 Resp: 33962
Ion Ratio Lower Upper
101 100
85 42.0 34.1 51.1
151 75.1 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#10
Methyl Iodide
Concen: 22.923 ug/l
RT: 4.169 min Scan# 439
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:142 Resp: 34640
Ion Ratio Lower Upper
142 100
127 40.4 31.9 47.9
141 14.4 11.4 17.2









#11
Tert butyl alcohol
Concen: 100.918 ug/l
RT: 5.052 min Scan# 589
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 59 Resp: 762
Ion Ratio Lower Upper
59 100
57 4.8 4.4 6.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#12
1,1-Dichloroethene
Concen: 22.681 ug/l
RT: 3.940 min Scan# 400
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 96 Resp: 32067
Ion Ratio Lower Upper
96 100
61 141.1 113.9 170.9
98 63.4 51.8 77.8





#13

Acrolein

Concen: 142.114 ug/l

RT: 3.799 min Scan# 31

Delta R.T. 0.006 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 56 Resp: 6129

Ion Ratio Lower Upper

56 100





55 69.6 54.0 81.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#14

Allyl chloride

Concen: 22.704 ug/l

RT: 4.563 min Scan# 506

Delta R.T. -0.001 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29




Tgt Ion: 41 Resp: 34335


Ion Ratio Lower Upper

41 100

39 64.9 53.4 80.2

76 53.2 42.4 63.6





#15

Acrylonitrile

Concen: 109.829 ug/l

RT: 5.263 min Scan# 61

Delta R.T. 0.011 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01





Tgt Ion: 53	Resp: 27989
Ion Ratio	Lower Upper
53 100	
52 81.0	66.0 99.0
51 34.3	28.3 42.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#16

Acetone

Concen: 102.761 ug/l




RT: 4.022 min Scan# 414


Delta R.T. 0.005 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29


Tgt Ion: 43	Resp: 24151
Ion Ratio	Lower Upper
43 100	
58 34.5	30.2 45.2





#17
Carbon Disulfide
Concen: 21.801 ug/l
RT: 4.275 min Scan# 456
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01




Tgt Ion: 76 Resp: 8768
Ion Ratio Lower Upper
76 100
78 9.8 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#18
Methyl Acetate
Concen: 23.849 ug/l
RT: 4.563 min Scan# 506
Delta R.T. -0.001 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29



Tgt Ion: 43 Resp: 15298
Ion Ratio Lower Upper
43 100
74 33.1 24.7 37.1









#19
Methyl tert-butyl Ether
Concen: 20.895 ug/l
RT: 5.316 min Scan# 611
Delta R.T. -0.001 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 73 Resp: 5914
Ion Ratio Lower Upper
73 100
57 21.1 16.1 24.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#20
Methylene Chloride
Concen: 20.463 ug/l
RT: 4.799 min Scan# 546
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 84 Resp: 56454
Ion Ratio Lower Upper
84 100
49 95.0 68.0 102.0
51 30.3 23.2 34.8
86 60.9 48.0 72.0





#21
trans-1,2-Dichloroethene
Concen: 22.293 ug/l
RT: 5.316 min Scan# 611
Delta R.T. 0.011 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01




Tgt Ion: 96 Resp: 3574
Ion Ratio Lower Upper
96 100
61 108.8 95.7 143.5
98 62.1 54.8 82.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#22
Diisopropyl ether
Concen: 22.463 ug/l
RT: 6.222 min Scan# 788
Delta R.T. 0.011 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Tgt Ion: 45 Resp: 75716
Ion Ratio Lower Upper
45 100
43 52.4 37.8 56.6
87 34.7 26.6 40.0
59 15.2 9.9 14.9#





#23
 Vinyl Acetate
 Concen: 103.199 ug/l m
 RT: 6.163 min Scan# 71
 Delta R.T. 0.011 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29




Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBS01



Tgt Ion: 43 Resp: 149290
 Ion Ratio Lower Upper
 43 100
 86 13.7 11.2 16.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022



#24
 1,1-Dichloroethane
 Concen: 22.942 ug/l
 RT: 6.116 min Scan# 770
 Delta R.T. -0.001 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29

Tgt Ion: 63 Resp: 57748
 Ion Ratio Lower Upper
 63 100
 98 9.3 4.6 13.8
 100 5.7 2.9 8.6





#25

2-Butanone

Concen: 93.424 ug/l

RT: 7.087 min Scan# 91

Delta R.T. 0.006 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 43 Resp: 2871

Ion Ratio Lower Upper

43 100





72 34.0 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#26

2,2-Dichloropropane

Concen: 19.516 ug/l

RT: 7.081 min Scan# 934

Delta R.T. 0.006 min

Lab File: VD074945.D




Acq: 16 Nov 2022 13:29


Tgt Ion: 77 Resp: 46165

Ion Ratio Lower Upper

77 100


97 24.2 12.5 37.5





#27
 cis-1,2-Dichloroethene
 Concen: 19.638 ug/l
 RT: 7.081 min Scan# 911
 Delta R.T. -0.000 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29




Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBS01



Tgt Ion: 96 Resp: 34929
 Ion Ratio Lower Upper
 96 100
 61 123.5 0.0 242.4
 98 64.5 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022



#28
 Bromochloromethane
 Concen: 19.917 ug/l
 RT: 7.428 min Scan# 993
 Delta R.T. 0.000 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29

Tgt Ion: 49 Resp: 16174
 Ion Ratio Lower Upper
 49 100
 129 1.0 0.0 5.0
 130 94.7 84.1 126.1





#29

Tetrahydrofuran

Concen: 95.353 ug/l

RT: 7.452 min Scan# 997

Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01





Tgt Ion:	42	Resp:	1622
Ion Ratio	Lower	Upper	
42	100		
72	53.6	45.8	68.6
71	50.2	44.6	66.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#30

Chloroform

Concen: 19.603 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29


Tgt Ion:	83	Resp:	54040
Ion Ratio	Lower	Upper	
83	100		
85	65.7	52.0	78.0





#31
Cyclohexane
Concen: 19.401 ug/l
RT: 7.881 min Scan# 1070
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 56 Resp: 4026
Ion Ratio Lower Upper
56 100
69 35.7 29.2 43.8
84 98.3 83.4 125.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#32
1,1,1-Trichloroethane
Concen: 20.024 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 97 Resp: 47573
Ion Ratio Lower Upper
97 100
99 64.9 53.3 79.9
61 40.3 30.2 45.4





#33

1,2-Dichloroethane-d4

Concen: 53.346 ug/l

RT: 8.234 min Scan# 1129

Delta R.T. 0.006 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 65 Resp: 5008

Ion Ratio Lower Upper

65 100





67 56.5 0.0 110.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1223

Delta R.T. 0.006 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29




Tgt Ion:114 Resp: 243860


Ion Ratio Lower Upper

114 100

63 17.2 0.0 33.2

88 14.2 0.0 28.2





#35

Dibromofluoromethane

Concen: 42.252 ug/l

RT: 7.810 min Scan# 1057

Delta R.T. 0.005 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 113 Resp: 63514





Ion	Ratio	Lower	Upper
113	100		
111	102.7	82.3	123.5
192	18.7	16.4	24.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#36

1,1-Dichloropropene

Concen: 18.191 ug/l

RT: 8.010 min Scan# 1092




Delta R.T. -0.001 min


Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Tgt Ion: 75 Resp: 40188

Ion	Ratio	Lower	Upper
75	100		
110	39.7	19.7	59.0
77	31.4	25.4	38.2





#37

Ethyl Acetate

Concen: 16.116 ug/l

RT: 7.169 min Scan# 948

Delta R.T. 0.005 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 43 Resp: 1194





Ion	Ratio	Lower	Upper
43	100		
61	16.2	12.2	18.4
70	10.4	9.6	14.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#38

Carbon Tetrachloride

Concen: 17.927 ug/l

RT: 7.999 min Scan# 1090




Delta R.T. 0.012 min


Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Tgt Ion:117 Resp: 37281

Ion	Ratio	Lower	Upper
117	100		
119	89.8	73.1	109.7
121	29.3	25.0	37.6





#39

Methylcyclohexane

Concen: 19.832 ug/l

RT: 9.275 min Scan# 1137

Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 83 Resp: 46938





Ion	Ratio	Lower	Upper
83	100		
55	64.8	49.0	73.4
98	51.1	41.7	62.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#40

Benzene

Concen: 18.892 ug/l

RT: 8.257 min Scan# 1134




Delta R.T. 0.005 min


Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Tgt Ion: 78 Resp: 120014


Ion	Ratio	Lower	Upper
78	100		
77	23.4	20.1	30.1





#41
Methacrylonitrile
Concen: 17.491 ug/l
RT: 7.410 min Scan# 989
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 41 Resp: 687
Ion Ratio Lower Upper
41 100
39 51.6 35.0 52.6
67 88.0 77.0 115.6
52 32.7 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#42
1,2-Dichloroethane
Concen: 17.961 ug/l
RT: 8.328 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 62 Resp: 27306
Ion Ratio Lower Upper
62 100
98 11.3 0.0 24.2





#43
Isopropyl Acetate
Concen: 17.503 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 43 Resp: 2285
Ion Ratio Lower Upper
43 100
61 40.5 30.9 46.3
87 18.0 14.3 21.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#44
Trichloroethene
Concen: 19.851 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:130 Resp: 33317
Ion Ratio Lower Upper
130 100
95 94.5 0.0 185.4





#45

1,2-Dichloropropane

Concen: 19.931 ug/l

RT: 9.304 min Scan# 1312

Delta R.T. -0.001 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 63 Resp: 28449

Ion Ratio Lower Upper

63 100





65 33.2 23.1 34.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#46

Dibromomethane

Concen: 19.619 ug/l

RT: 9.393 min Scan# 1327

Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29




Tgt Ion: 93 Resp: 15683


Ion Ratio Lower Upper

93 100

95 81.9 68.2 102.2

174 96.1 87.4 131.0









#47
Bromodichloromethane
Concen: 19.586 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 83 Resp: 3885
Ion Ratio Lower Upper
83 100
85 63.7 48.8 73.2
127 9.0 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#48
Methyl methacrylate
Concen: 18.391 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 41 Resp: 10637
Ion Ratio Lower Upper
41 100
69 111.7 93.0 139.4
39 56.0 44.6 66.8





#49

1,4-Dioxane

Concen: 373.534 ug/l

RT: 9.381 min Scan# 1325

Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 88 Resp: 2930

Ion Ratio Lower Upper

88 100

43 26.1 24.1 36.1





58 61.3 43.1 64.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#50

Toluene-d8

Concen: 46.971 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VD074945.D




Acq: 16 Nov 2022 13:29


Tgt Ion: 98 Resp: 232588

Ion Ratio Lower Upper

98 100

100 65.4 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 94.567 ug/l
RT: 10.163 min Scan# 1458
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 43 Resp: 58864
Ion Ratio Lower Upper
43 100
58 44.9 37.8 56.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#52
Toluene
Concen: 20.307 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 92 Resp: 77132
Ion Ratio Lower Upper
92 100
91 181.8 139.0 208.4





#53
t-1,3-Dichloropropene
Concen: 19.426 ug/l
RT: 10.557 min Scan# 1525
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01




Tgt Ion: 75 Resp: 3404
Ion Ratio Lower Upper
75 100
77 31.8 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#54
cis-1,3-Dichloropropene
Concen: 19.881 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Tgt Ion: 75 Resp: 42730
Ion Ratio Lower Upper
75 100
77 30.2 24.7 37.1
39 30.9 25.4 38.2





#55
1,1,2-Trichloroethane
Concen: 20.120 ug/l
RT: 10.734 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion	Ratio	Lower	Upper
97	100		
83	83.6	64.8	97.2
85	55.4	44.8	67.2
99	60.7	49.4	74.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#56
Ethyl methacrylate
Concen: 18.481 ug/l
RT: 10.598 min Scan# 1532
Delta R.T. -0.001 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion	Ratio	Lower	Upper
69	100		
41	50.8	38.6	58.0
39	25.6	19.9	29.9





#57

1,3-Dichloropropane

Concen: 19.353 ug/l

RT: 10.881 min Scan# 1580

Delta R.T. -0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01

Tgt Ion: 76 Resp: 3458

Ion Ratio Lower Upper

76 100





78 34.1 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#58

2-Chloroethyl Vinyl ether

Concen: 96.462 ug/l

RT: 9.875 min Scan# 1409

Delta R.T. 0.006 min

Lab File: VD074945.D




Acq: 16 Nov 2022 13:29


Tgt Ion: 63 Resp: 38964

Ion Ratio Lower Upper

63 100


106 31.7 26.5 39.7





#59
2-Hexanone
Concen: 95.277 ug/l
RT: 10.922 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01



Tgt Ion: 43 Resp: 4134
Ion Ratio Lower Upper
43 100
58 62.4 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#60
Dibromochloromethane
Concen: 19.911 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:129 Resp: 26255
Ion Ratio Lower Upper
129 100
127 76.2 39.1 117.2









#61
1,2-Dibromoethane
Concen: 19.983 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:107 Resp: 2077
Ion Ratio Lower Upper
107 100
109 92.5 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#62
4-Bromofluorobenzene
Concen: 50.562 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 95 Resp: 76880
Ion Ratio Lower Upper
95 100
174 87.9 0.0 197.0
176 81.5 0.0 182.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01




Tgt Ion:117 Resp: 22041
Ion Ratio Lower Upper
117 100
82 50.0 41.6 62.4
119 32.6 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#64
Tetrachloroethene
Concen: 18.907 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29



Tgt Ion:164 Resp: 27941
Ion Ratio Lower Upper
164 100
166 116.4 107.8 161.8
129 84.8 77.4 116.2
131 85.9 69.5 104.3









#65
Chlorobenzene
Concen: 19.528 ug/l
RT: 11.610 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:112 Resp: 8420
Ion Ratio Lower Upper
112 100
114 29.7 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 19.159 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:131 Resp: 28945
Ion Ratio Lower Upper
131 100
133 95.2 46.6 139.8
119 65.0 32.5 97.4









#67
Ethyl Benzene
Concen: 19.462 ug/l
RT: 11.687 min Scan# 1717
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 91 Resp: 146199
Ion Ratio Lower Upper
91 100
106 31.5 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#68
m/p-Xylenes
Concen: 39.455 ug/l
RT: 11.792 min Scan# 1735
Delta R.T. -0.001 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:106 Resp: 118174
Ion Ratio Lower Upper
106 100
91 191.9 156.2 234.4









#69
o-Xylene
Concen: 19.370 ug/l
RT: 12.128 min Scan# 1791
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:106 Resp: 53370
Ion Ratio Lower Upper
106 100
91 202.0 104.5 313.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#70
Styrene
Concen: 19.320 ug/l
RT: 12.140 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:104 Resp: 91261
Ion Ratio Lower Upper
104 100
78 43.8 35.7 53.5
103 56.3 43.7 65.5









#71
Bromoform
Concen: 18.039 ug/l
RT: 12.304 min Scan# 1822
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:173 Resp: 14104
Ion Ratio Lower Upper
173 100
175 48.5 23.1 69.3
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion:152 Resp: 102758
Ion Ratio Lower Upper
152 100
115 56.4 26.8 80.4
150 165.1 0.0 348.2





#73
Isopropylbenzene
Concen: 20.303 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBSD01




Tgt Ion:105 Resp: 143598
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#74
N-ethyl acetate
Concen: 18.115 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Tgt Ion: 43 Resp: 21080
Ion Ratio Lower Upper
43 100
70 54.9 43.8 65.6
55 27.9 22.0 33.0
61 29.7 23.9 35.9





#75
1,1,2,2-Tetrachloroethane
Concen: 19.752 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 83 Resp: 2471
Ion Ratio Lower Upper
83 100
131 9.3 4.7 14.1
85 63.1 33.1 99.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#76
1,2,3-Trichloropropane
Concen: 19.803 ug/l m
RT: 12.728 min Scan# 1894
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 75 Resp: 17020
Ion Ratio Lower Upper
75 100
77 0.0 185.8 557.5#









#77
Bromobenzene
Concen: 19.421 ug/l
RT: 12.704 min Scan# 1889
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:156 Resp: 32145
Ion Ratio Lower Upper
156 100
77 166.4 77.7 233.1
158 98.6 49.4 148.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#78
n-propylbenzene
Concen: 20.568 ug/l
RT: 12.769 min Scan# 1901
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion: 91 Resp: 178820
Ion Ratio Lower Upper
91 100
120 23.0 11.9 35.7









#79
2-Chlorotoluene
Concen: 20.540 ug/l
RT: 12.851 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 91 Resp: 96218
Ion Ratio Lower Upper
91 100
126 36.8 18.3 54.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#80
1,3,5-Trimethylbenzene
Concen: 20.108 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. 0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion:105 Resp: 118503
Ion Ratio Lower Upper
105 100
120 49.3 25.7 77.0





#81
trans-1,4-Dichloro-2-butene
Concen: 17.763 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01



Tgt Ion: 75 Resp: 6200
Ion Ratio Lower Upper
75 100
53 82.2 63.8 95.6
89 51.2 38.2 57.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#82
4-Chlorotoluene
Concen: 20.440 ug/l
RT: 12.951 min Scan# 1932
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 91 Resp: 99712
Ion Ratio Lower Upper
91 100
126 35.9 18.8 56.4





#83
tert-Butylbenzene
Concen: 19.609 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01




Tgt Ion:119 Resp: 100274
Ion Ratio Lower Upper
119 100
91 62.2 30.0 90.0
134 27.1 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#84
1,2,4-Trimethylbenzene
Concen: 20.228 ug/l
RT: 13.216 min Scan# 1977
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Tgt Ion:105 Resp: 117901
Ion Ratio Lower Upper
105 100
120 46.1 22.9 68.7





#85
 sec-Butylbenzene
 Concen: 20.559 ug/l
 RT: 13.345 min Scan# 1999
 Delta R.T. -0.001 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29




Instrument :
 MSVOA_D
 ClientSampleId :
 VD1116SBS01




Tgt Ion:105 Resp: 159910
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.0 30.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 11/17/2022
 Supervised By :Mahesh Dadoda 11/17/2022




#86
 p-Isopropyltoluene
 Concen: 20.346 ug/l
 RT: 13.463 min Scan# 2019
 Delta R.T. -0.000 min
 Lab File: VD074945.D
 Acq: 16 Nov 2022 13:29



Tgt Ion:119 Resp: 131595
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.5 40.4
 91 21.4 10.8 32.4





#87

1,3-Dichlorobenzene

Concen: 19.457 ug/l

RT: 13.457 min Scan# 2018

Delta R.T. 0.000 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Instrument :

MSVOA_D

ClientSampleId :

VD1116SBS01





Tgt Ion:146 Resp: 66989
 Ion Ratio Lower Upper
 146 100
 111 37.7 18.3 54.8
 148 64.5 31.8 95.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/17/2022

Supervised By :Mahesh Dadoda 11/17/2022



#88

1,4-Dichlorobenzene

Concen: 19.203 ug/l




RT: 13.539 min Scan# 2032


Delta R.T. -0.001 min

Lab File: VD074945.D

Acq: 16 Nov 2022 13:29

Tgt Ion:146 Resp: 65619
 Ion Ratio Lower Upper
 146 100
 111 39.7 17.7 53.1
 148 64.9 32.3 96.9









#89
n-Butylbenzene
Concen: 20.419 ug/l
RT: 13.792 min Scan# 2074
Delta R.T. 0.005 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion: 91 Resp: 122614
Ion Ratio Lower Upper
91 100
92 52.2 26.9 80.7
134 25.6 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#90
Hexachloroethane
Concen: 20.821 ug/l
RT: 14.057 min Scan# 2120
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:117 Resp: 24119
Ion Ratio Lower Upper
117 100
201 68.2 38.0 114.0









#91
1,2-Dichlorobenzene
Concen: 19.876 ug/l
RT: 13.834 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:146 Resp: 5819
Ion Ratio Lower Upper
146 100
111 38.1 18.9 56.5
148 63.8 31.9 95.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.987 ug/l
RT: 14.445 min Scan# 2186
Delta R.T. -0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29


Tgt Ion: 75 Resp: 2930
Ion Ratio Lower Upper
75 100
155 106.7 51.2 153.8
157 124.3 68.2 204.5





#93
1,2,4-Trichlorobenzene
Concen: 18.163 ug/l
RT: 15.104 min Scan# 21
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29




Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01




Tgt Ion:180 Resp: 32370
Ion Ratio Lower Upper
180 100
182 98.8 46.9 140.6
145 30.6 14.1 42.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022




#94
Hexachlorobutadiene
Concen: 20.305 ug/l
RT: 15.210 min Scan# 2316
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29



Tgt Ion:225 Resp: 18666
Ion Ratio Lower Upper
225 100
223 63.5 31.9 95.5
227 64.8 32.6 97.8









#95
Naphthalene
Concen: 17.789 ug/l
RT: 15.333 min Scan# 21
Delta R.T. -0.001 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Instrument :
MSVOA_D
ClientSampleId :
VD1116SBS01

Tgt Ion:128 Resp: 5884
Ion Ratio Lower Upper
128 100
127 13.9 10.6 16.0
129 11.5 8.3 12.5

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/17/2022
Supervised By :Mahesh Dadoda 11/17/2022



#96
1,2,3-Trichlorobenzene
Concen: 19.116 ug/l
RT: 15.528 min Scan# 2370
Delta R.T. 0.006 min
Lab File: VD074945.D
Acq: 16 Nov 2022 13:29

Tgt Ion:180 Resp: 29461
Ion Ratio Lower Upper
180 100
182 90.4 46.7 140.1
145 30.3 14.9 44.7

