

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
 Data File : VD074954.D
 Acq On : 16 Nov 2022 17:42
 Operator : VA/SY
 Sample : N5642-09
 Misc : 6.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LEFT-SIDE-FLATS-29

Quant Time: Nov 17 00:21:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	138303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	253211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	234124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	99384	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66506	75.377	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.760%
35) Dibromofluoromethane	7.810	113	83444	53.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.920%
50) Toluene-d8	10.275	98	310142	60.320	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	120.640%
62) 4-Bromofluorobenzene	12.575	95	85776	54.330	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.660%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111622\
Data File : VD074954.D
Acq On : 16 Nov 2022 17:42
Operator : VA/SY
Sample : N5642-09
Misc : 6.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-29

Quant Time: Nov 17 00:21:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.881 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42




Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-29

Tgt Ion:168 Resp: 138303
Ion Ratio Lower Upper
168 100
99 50.1 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 75.377 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42

Tgt Ion: 65 Resp: 66506
Ion Ratio Lower Upper
65 100
67 55.5 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074954.D

Acq: 16 Nov 2022 17:42

Instrument :

MSVOA_D

ClientSampleId :

LEFT-SIDE-FLATS-29





Tgt Ion:114 Resp: 253211

Ion Ratio Lower Upper

114 100

63 15.5 0.0 33.2

88 14.5 0.0 28.2



#35

Dibromofluoromethane

Concen: 53.460 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.005 min

Lab File: VD074954.D

Acq: 16 Nov 2022 17:42




Tgt Ion:113 Resp: 83444


Ion Ratio Lower Upper

113 100

111 103.5 82.3 123.5

192 19.1 16.4 24.6









#50
Toluene-d8
Concen: 60.320 ug/l
RT: 10.275 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42




Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-29

Tgt Ion: 98 Resp: 310142
Ion Ratio Lower Upper
98 100
100 64.0 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 54.330 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42

Tgt Ion: 95 Resp: 85776
Ion Ratio Lower Upper
95 100
174 93.8 0.0 197.0
176 91.1 0.0 182.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42

Instrument :
MSVOA_D
ClientSampleId :
LEFT-SIDE-FLATS-29

Tgt Ion:117 Resp: 234124
Ion Ratio Lower Upper
117 100
82 48.8 41.6 62.4
119 31.7 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.522 min Scan# 2029
Delta R.T. 0.006 min
Lab File: VD074954.D
Acq: 16 Nov 2022 17:42

Tgt Ion:152 Resp: 99384
Ion Ratio Lower Upper
152 100
115 55.3 26.8 80.4
150 158.6 0.0 348.2

