




Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111723\
 Data File : VD077603.D
 Acq On : 17 Nov 2023 10:35
 Operator : JC/SY
 Sample : 05437-01
 Misc : 6.25G/10ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EY1E2

Manual IntegrationsAPPROVED

Quant Time: Nov 18 00:26:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 18 00:25:31 2023
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 11/20/2023
 Supervised By :Semsettin Yesilyurt 11/21/2023



TIC: VD077603.D\data.ms

(16) Methylene chloride (T)

4.799min (-0.006) 4.73 ug/L

response 4282

Ion	Exp%	Act%
84.00	100.00	100.00
86.00	66.50	46.47#
49.00	140.50	131.48
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111723\
Data File : VD077603.D
Acq On : 17 Nov 2023 10:35
Operator : JC/SY
Sample : 05437-01
Misc : 6.25G/10ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EY1E2

Manual IntegrationsAPPROVED

Quant Time: Nov 18 00:26:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Nov 18 00:25:31 2023
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 11/20/2023
Supervised By :Semsettin Yesilyurt 11/21/2023



TIC: VD077603.D\data.ms

(16) Methylene chloride (T)

4.799min (-0.006) 7.74 ug/L m

response 7014

Ion	Exp%	Act%
84.00	100.00	100.00
86.00	66.50	46.47#
49.00	140.50	131.48
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111723\
 Data File : VD077603.D
 Acq On : 17 Nov 2023 10:35
 Operator : JC/SY
 Sample : 05437-01
 Misc : 6.25G/10ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EY1E2

Manual IntegrationsAPPROVED

Quant Time: Nov 18 00:26:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 18 00:25:31 2023
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 11/20/2023
 Supervised By :Semsettin Yesilyurt 11/21/2023

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.781	114		53641	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117		50277	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152		22910	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.281	65		20489	15.015	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	60.080%		
7) Chloroethane-d5	2.811	69		18886	17.618	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	70.480%		
11) 1,1-Dichloroethene-d2	3.922	65		6492	15.386	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	61.560%		
21) 2-Butanone-d5	6.993	46		22148	93.748	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	187.500%#		
24) Chloroform-d	7.569	84		31983	20.640	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	82.560%		
26) 1,2-Dichloroethane-d4	8.234	65		22288	27.041	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	108.160%		
32) Benzene-d6	8.204	84		60005	18.389	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	73.560%		
36) 1,2-Dichloropropane-d6	9.210	67		22540	21.206	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	84.840%		
41) Toluene-d8	10.269	98		51905	17.603	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	70.400%		
43) trans-1,3-Dichloroprop...	10.522	79		10453	22.809	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery =	91.240%		
47) 2-Hexanone-d5	10.881	63		16500	86.777	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	173.560%#		
56) 1,1,2,2-Tetrachloroeth...	12.651	84		25336	35.977	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	143.920%#		
66) 1,2-Dichlorobenzene-d4	13.816	152		17916	21.982	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	87.920%		
Target Compounds							
13) Acetone	4.022	43		6014	24.217	ug/L	89
16) Methylene chloride	4.799	84		7014m	7.745	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111723\
Data File : VD077603.D
Acq On : 17 Nov 2023 10:35
Operator : JC/SY
Sample : 05437-01
Misc : 6.25G/10ml/MSVOA_D/SoIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EY1E2

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 11/20/2023
Supervised By :Semsettin Yesilyurt 11/21/2023

Quant Time: Nov 18 00:26:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Nov 18 00:25:31 2023
Response via : Initial Calibration

