

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
 Data File : VD080042.D
 Acq On : 17 Nov 2024 15:19
 Operator : RP/MD
 Sample : P4810-02
 Misc : 5.02G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Nov 18 04:24:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	124	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.793	114	150	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.581	117	420	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.975	65	54	42.297	ug/l	-0.25
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.600%
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.275	98	132	35.082	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	70.160%
62) 4-Bromofluorobenzene	12.569	95	368	298.106	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	596.220%#
Target Compounds						Qvalue
20) Methylene Chloride	4.810	84	4114	2769.267	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
Data File : VD080042.D
Acq On : 17 Nov 2024 15:19
Operator : RP/MD
Sample : P4810-02
Misc : 5.02G/5ml/MSVOA_D/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Nov 18 04:24:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

