

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
 Data File : VD080047.D
 Acq On : 17 Nov 2024 17:34
 Operator : RP/MD
 Sample : P4833-07
 Misc : 4.33G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2-VOC

Quant Time: Nov 18 04:26:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	12870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	12566	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	4240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	22943	549.531	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 1099.060%#	
35) Dibromofluoromethane	7.810	113	13983	163.250	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 326.500%#	
50) Toluene-d8	10.275	98	12728	39.426	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 78.860%	
62) 4-Bromofluorobenzene	12.575	95	5258	49.643	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.280%	
Target Compounds						
16) Acetone	4.022	43	1772	195.615	ug/l	# 24

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
Data File : VD080047.D
Acq On : 17 Nov 2024 17:34
Operator : RP/MD
Sample : P4833-07
Misc : 4.33G/5ml/MSVOA_D/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2-VOC

Quant Time: Nov 18 04:26:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

