

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
 Data File : VD080049.D
 Acq On : 17 Nov 2024 18:28
 Operator : RP/MD
 Sample : P4825-01
 Misc : 5.01G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCK-PILE

Quant Time: Nov 18 04:27:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	15211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	33408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	37593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	18656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	20387	130.175	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	260.360%#
35) Dibromofluoromethane	7.810	113	16552	74.444	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	148.880%#
50) Toluene-d8	10.269	98	35904	42.845	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.680%
62) 4-Bromofluorobenzene	12.575	95	15186	55.234	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111724\
Data File : VD080049.D
Acq On : 17 Nov 2024 18:28
Operator : RP/MD
Sample : P4825-01
Misc : 5.01G/5ml/MSVOA_D/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCK-PILE

Quant Time: Nov 18 04:27:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

