

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD111819\
 Data File : VD064270.D
 Acq On : 18 Nov 2019 17:55
 Operator : VA/SY
 Sample : K5839-06RE
 Misc : 5.18G/5.00ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PHIPPS-BH-NW-03RE

Quant Time: Nov 19 05:55:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 12 02:43:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	514116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	780210	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	720533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	373708	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.34	65	293411	69.39	ug/l	0.00
Spiked Amount			Recovery	=	138.78%	
35) Dibromofluoromethane	7.92	113	252739	56.43	ug/l	0.00
Spiked Amount			Recovery	=	112.86%	
50) Toluene-d8	10.34	98	921460	54.54	ug/l	0.00
Spiked Amount			Recovery	=	109.08%	
62) 4-Bromofluorobenzene	12.64	95	325164	55.58	ug/l	0.00
Spiked Amount			Recovery	=	111.16%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD111819\
Data File : VD064270.D
Acq On : 18 Nov 2019 17:55
Operator : VA/SY
Sample : K5839-06RE
Misc : 5.18G/5.00ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PHIPPS-BH-NW-03RE

Quant Time: Nov 19 05:55:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 12 02:43:24 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. -0.00 min

Lab File: VD064270.D

Acq: 18 Nov 2019 17:55

Instrument :

MSVOA_D

ClientSampleId :


PHIPPS-BH-NW-03RE

Tot Ion:168 Resp: 514116

Ion Ratio Lower Upper




168 100

99 50.1 41.6 62.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 69.392 ug/l

RT: 8.34 min Scan# 1091

Delta R.T. -0.00 min

Lab File: VD064270.D


Acq: 18 Nov 2019 17:55

Tgt Ion: 65 Resp: 293411

Ion Ratio Lower Upper



65 100


67 51.1 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. -0.00 min

Lab File: VD064270.D

Acq: 18 Nov 2019 17:55

Instrument :

MSVOA_D

ClientSampleId :

PHIPPS-BH-NW-03RE


Tot Ion:114 Resp: 780210

Ion Ratio Lower Upper

114 100

63 19.0 0.0 40.0




88 15.0 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VDC

Ion 87.90 (87.60 to 88.60): VDC



#35

Dibromofluoromethane

Concen: 56.433 ug/l

RT: 7.92 min Scan# 1020

Delta R.T. -0.00 min

Lab File: VD064270.D

Acq: 18 Nov 2019 17:55


Tgt Ion:113 Resp: 252739

Ion Ratio Lower Upper

113 100

111 102.9 81.5 122.3



192 23.5 18.8 28.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 54.541 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. -0.00 min

Lab File: VD064270.D

Acq: 18 Nov 2019 17:55

Instrument :

MSVOA_D

ClientSampleId :





PHIPPS-BH-NW-03RE

Tot Ion: 98 Resp: 921460

Ion Ratio Lower Upper

98 100

100 64.7 52.6 78.8



#62

4-Bromofluorobenzene

Concen: 55.576 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. -0.00 min

Lab File: VD064270.D

Acq: 18 Nov 2019 17:55




Tgt Ion: 95 Resp: 325164


Ion Ratio Lower Upper

95 100

174 95.9 0.0 183.6

176 93.6 0.0 180.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. -0.00 min
Lab File: VD064270.D
Acq: 18 Nov 2019 17:55

Instrument :
MSVOA_D
ClientSampleId :
PHIPPS-BH-NW-03RE

Tot Ion	Ratio	Lower	Upper
117	100		
82	53.2	42.8	64.2
119	30.9	25.6	38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.01 min
Lab File: VD064270.D
Acq: 18 Nov 2019 17:55

Tgt Ion	Ratio	Lower	Upper
152	100		
115	54.2	27.4	82.2
150	156.2	0.0	351.6

