

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111822\
 Data File : VD074993.D
 Acq On : 18 Nov 2022 16:58
 Operator : VA/SY
 Sample : N5695-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-1-111722

Quant Time: Nov 18 22:06:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.88
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.78
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.58
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.52

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111822\
Data File : VD074993.D
Acq On : 18 Nov 2022 16:58
Operator : VA/SY
Sample : N5695-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722

Quant Time: Nov 18 22:06:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 0.000 ug/l
Expected RT: 7.88 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722


Tgt Ion: 168
Sig Exp Ratio
168 100
99 51.1




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.23 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Tgt Ion: 65
Sig Exp Ratio
65 100
67 55.1







#34
1,4-Difluorobenzene
Concen: 0.000 ug/l
Expected RT: 8.78 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722


Tgt Ion: 114
Sig Exp Ratio
114 100
63 16.6
88 14.1




#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 7.80 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Tgt Ion: 113
Sig Exp Ratio
113 100
111 102.9
192 20.5







#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.27 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722


Tgt Ion: 98
Sig Exp Ratio
98 100
100 64.2




#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.57 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Tgt Ion: 95
Sig Exp Ratio
95 100
174 98.5
176 91.3







#63
Chlorobenzene-d5
Concen: 0.000 ug/l
Expected RT: 11.58 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Instrument :
MSVOA_D
ClientSampleId :
HD-1-111722

Tgt Ion: 117
Sig Exp Ratio
117 100
82 52.0
119 31.5



#72
1,4-Dichlorobenzene-d4
Concen: 0.000 ug/l
Expected RT: 13.52 min

Lab File: VD074993.D
Acq: 18 Nov 2022 16:58

Tgt Ion: 152
Sig Exp Ratio
152 100
115 53.6
150 174.1

