

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111822\
 Data File : VD074995.D
 Acq On : 18 Nov 2022 17:52
 Operator : VA/SY
 Sample : N5697-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-03-111722

Quant Time: Nov 18 22:06:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 14 00:51:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.887	168	2954	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.775	114	6142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	6578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	2918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.263	98	55	0.441	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.880%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111822\
Data File : VD074995.D
Acq On : 18 Nov 2022 17:52
Operator : VA/SY
Sample : N5697-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-03-111722

Quant Time: Nov 18 22:06:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111122S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 14 00:51:58 2022
Response via : Initial Calibration

