

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111823\
 Data File : VD077625.D
 Acq On : 18 Nov 2023 12:37
 Operator : JC/SY
 Sample : 05416-24
 Misc : 5.75G/10ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0AH7

Quant Time: Nov 19 23:23:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Nov 19 23:22:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	425265	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	366736	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	139497	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	169105	15.632	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.520%
7) Chloroethane-d5	2.805	69	156265	18.387	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.560%
11) 1,1-Dichloroethene-d2	3.928	65	53037	15.855	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.400%
21) 2-Butanone-d5	6.987	46	58117	31.029	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.060%
24) Chloroform-d	7.569	84	241531	19.661	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.640%
26) 1,2-Dichloroethane-d4	8.228	65	119295	18.256	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.040%
32) Benzene-d6	8.204	84	479608	20.150	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	9.210	67	156390	20.171	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.680%
41) Toluene-d8	10.269	98	410540	19.087	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.360%
43) trans-1,3-Dichloroprop...	10.528	79	57633	17.240	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.960%
47) 2-Hexanone-d5	10.875	63	43230	31.169	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.340%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	90523	17.622	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	70.480%
66) 1,2-Dichlorobenzene-d4	13.816	152	103687	20.894	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.560%
Target Compounds						
16) Methylene chloride	4.805	84	61357	8.545	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111823\
Data File : VD077625.D
Acq On : 18 Nov 2023 12:37
Operator : JC/SY
Sample : 05416-24
Misc : 5.75G/10ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0AH7

Quant Time: Nov 19 23:23:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM110923SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Nov 19 23:22:21 2023
Response via : Initial Calibration

