

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
 Data File : VD080062.D
 Acq On : 18 Nov 2024 12:08
 Operator : RP/MD
 Sample : P4825-01RE
 Misc : 4.99G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCK-PILERE

Quant Time: Nov 18 14:29:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	8236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	15761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	18558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	8748	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	11256	132.740	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	265.480%#
35) Dibromofluoromethane	7.798	113	9223	87.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	175.860%#
50) Toluene-d8	10.269	98	21701	54.891	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.780%
62) 4-Bromofluorobenzene	12.569	95	8248	63.589	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	127.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
Data File : VD080062.D
Acq On : 18 Nov 2024 12:08
Operator : RP/MD
Sample : P4825-01RE
Misc : 4.99G/5ml/MSVOA_D/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCK-PILERE

Quant Time: Nov 18 14:29:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

